



₋umen CVC

Two-Lumen Catheter: 7 Fr. (2.5 mm OD) x 30 cm Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One End -"J" Tip on Other) with Arrow Advancer with ECG Mark Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW Syringe: 5 mL Luer-Slip

Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm

Dust Cap: Non-Vented
SecondSite™ Adjustable Hub: Fastener
SecondSite™ Adjustable Hub: Catheter Clamp

All components are CE 2797 unless otherwise noted.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU Not made with natural rubber latex. Fluid path components are non-pyrogenic.

de	el h	I is		
it	V (I	no		
pl	pt ro	ru		
Sr	sk s	es		
sv tr uk				
\bigwedge	MD	$\bigcap_{\mathbf{i}}$		





6 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (16 Ga.) Proximal (16 Ga.)	0.57 0.62	1917 2000	
* Priming volumes	are approximate	e and are done with	nout accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.





Teleflex Medical EC REP IDA Business and Technology Park Dublin Road, Athlone 2797 Co. Westmeath, Ireland

All components are CE 2797 unless otherwise noted.

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA

> Product of xxx Packaged in xxx



(01)XXXXXXXXXXXXXXX (17)YYMMDD

(11)YYMMDD (10)Lot Number

LBL055593 R03 (2024-10)















