



REF CS-16402

(IPNXXXXXX)

30 cm catheter length of the spring-wire guide spring-wire

Pediatric Two-Lumen CVC for Femoral Vein Insertion

- Two-Lumen Catheter: 4 Fr. $(1.5 \text{ mm OD}) \times 30 \text{ cm}$ Spring-Wire Guide: .018" (0.46 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One End "J" Tip on Other)
- Catheter: 22 Ga. x 1-3/4" (4.45 cm) Radiopaque over 25 Ga. RW Introducer Needle Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW

- Syringe: 5 mL Luer-Slip Tissue Dilator: 5 Fr. (1.8 mm) x 7.5 cm
- Tissue Dilator: 5 Fr. (1.8 mm) x 4.1 cm Dust Cap: Non-Vented

- SecondSite™ Adjustable Hub: Catheter Clamp
 SecondSite™ Adjustable Hub: Fastener
 All components are CE 2797 unless otherwise noted.

Rx	10	ı۱	1
----	----	----	---

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

California Prop. 65 ↑ WARNING: Cancer and Reproductive Harm.

www.P65Warnings.ca.gov

Not made with natural rubber latex.

Fluid path components are non-pyrogenic

Lumen Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (22 Ga.)	0.33	224	
Proximal (22 Ga.)	0.38	125	

^{*} Priming volumes are approximate and are done without accessories





EU Authorized Representative and Importer:



Teleflex Medical IDA Business and Technology Park
Dublin Road, Athlone
7797 Co. Westmeath, Ireland

All components are CE 2797 unless otherwise noted.

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA

Product of xxx Packaged in xxx

Neleflex®

LBL056223 R02 (2024-08)



LOT LotNumber

√ YYYY-MM-DD



M YYYY-MM-DD














(01)XXXXXXXXXXXXXXX (17)YYMMDD (11)YYMMDD











1018 inch dia. spring-wire quide



CAS # 7440-48-4

[†] Flow rate values are approximate and are determined using deionized water at 100 cm head height.