
**ARRC** 

Product of XXX





(IPNXXXXXX)

16 cm catheter length

.032 inch dia. spring-wire guide

## Arrowg+ard Blue® Three-Lumen CVC

- Arrowg+ard Blue® Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 16 cm
- Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow
- Advancer Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW Pressure Transduction Probe
- Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm

- Dust Cap: Non-Vented
  SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp

All components are CE 2797 unless otherwise noted.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F). Fluid path components are non-pyrogenic.

Contraindications: The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.







Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (16 Ga.)	0.42	2656	
Medial (18 Ga.)	0.38	1353	
Proximal (18 Ga.)	0.41	1477	

<sup>\*</sup> Priming volumes are approximate and are done without accessories.



Representative and Importer: Teleflex Medical

EC REP IDA Business and Technology Park Dublin Road, Athlone Co. Westmeath, Ireland

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA











(01)XXXXXXXXXXXXXXX (11)YYMMDD (10)LotNumber

LBL056441 R02 (2024-02)



REF CS-22703-E

16 cm catheter length

.032 inch dia. spring-wire guide





 $<sup>\</sup>uparrow$  Flow rate values are approximate and are determined using deionized water at 100 cm head height.