
ARRC

Product of xxx





(IPNXXXXXX)

20 cm catheter length

.032 inch dia. spring-wire quide

Arrowg+ard Blue Plus® Two-Lumen CVC

- Arrowg+ard Blue Plus®1 Two-Lumen Catheter: 7 Fr. (2.5 mm OD) x 20 cm
- Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer with ECG Mark
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW Pressure Transduction Probe
- Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm

- Dust Cap: Non-Vented
 SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp

Licensed under US Patent No. 7,329,412

All components are CE 2797 unless otherwise noted.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F). Fluid path components are non-pyrogenic.

Contraindications: The Arrowg+ard Blue Plus antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.





Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (16 Ga.)	0.42	2615	
Proximal (16 Ga.)	0.45	2324	

Priming volumes are approximate and are done without accessories.

EU Authorized Representative and Importer:

Teleflex Medical EC REP IDA Business and Technology Park Dublin Road, Athlone Co. Westmeath, Ireland

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL056470 R01 (2024-03)









(01)XXXXXXXXXXXXXXX (11)YYMMDD (10)LotNumber



REF CS-46702-E

20 cm catheter length

.032 inch dia. spring-wire guide





[†] Flow rate values are approximate and are determined using deionized water at 100 cm head height.