



REF PR-05552-HPX (IPNXXXXXX)

5 Fr. 2 Lumen 55 cm catheter length .018 inch dia. spring-wire guide

# Pressure Injectable Two-Lumen PICC with 130 cm Wire

**Contents:**

- 1: Two Lumen Catheter: 5 Fr. (1.81 mm OD) x 55 cm, Pressure Injectable, Blue FlexTip® Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: GlideThru™ Peel-Away Sheath: 5 Fr. x 4" (10 cm) Radiopaque over 5 Fr. Dilator Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on Both Ends) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Spring-Wire Guide, Marked: .018" (0.46 mm) dia. x 51-3/16" (130 cm)(Straight Soft Tip on One End - Straight Stiff Tip on Other) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Injection Needle: SafetyGlide™ 25 Ga. x 5/8" (1.60 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Injection Needle: SafetyGlide™ 21 Ga. x 1-1/2" (3.81 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 2: Syringe: 10 mL Luer-Lock Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Syringe: 3 mL Luer-Lock Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: 10 mL Luer-Lock Pre-Filled Saline Syringe Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: 3 mL Applicator 2% CHG and 70% IPA Chloraprep®2 One-Step Solution with Hi-Lite Orange™ Tint Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 2: Dust Cap: Non-Vented Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: SecondSite™ Adjustable Hub: Fastener Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: SecondSite™ Adjustable Hub: Catheter Clamp Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: SharpsAway® II Locking Disposal Cup Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: SharpsAway® Disposal Cup Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Catheter Trimmer Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Towel: 24" x 36" (61 cm x 91 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Safety Scalpel: #11 Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Patient ID Card Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Chart Sticker Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Patient Information Booklet Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Medication Label: 1% Lidocaine Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Medication Label: 0.9% Saline Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Checklist/CLIP Sheet Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 2: Paper Tape Measure Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Tourniquet Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Dressing: I.V. Advanced 8.5 cm x 11.5 cm Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Dressing: BIOPATCH®3 Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Dressing: STATLOCK®4 PICC Plus Catheter Stabilization Device Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX
- 1: Tape: Steri-Strip®5 Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXX

**Rx only**

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.**

**California Prop. 65**

**WARNING:** Cancer and Reproductive Harm. www.P65Warnings.ca.gov

**Not made with natural rubber latex.**

Store between 20 – 25°C (68 – 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

**Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.**



Lumen

	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.48	356	2934	4
Proximal (18 Ga.)	0.51	356	3317	4

\* Priming volumes are approximate and are done without accessories.  
 † Flow rate values are approximate and are determined using deionized water at 100 cm head height.  
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.  
 \*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



LOT Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXXXX  
(17)YYMMDD  
(11)YYMMDD  
(10)Lot Number

Set Product of xxx  
Packaged in xxx

Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA



LBL056514 R06 (2024-12)



REF PR-05552-HPX

5 Fr. 2 Lumen 55 cm catheter length .018 inch dia. spring-wire guide



# Pressure Injectable Two-Lumen PICC with 130 cm Wire