
ARRC

Product of XXX





(IPNXXXXXX)

20 cm catheter length

.032 inch dia. spring-wire guide

Arrowg+ard Blue® Two-Lumen CVC

- Arrowg+ard Blue® Two-Lumen Catheter: 8 Fr. (2.8 mm OD) x 20 cm
- Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer with ECG Mark
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW
- Pressure Transduction Probe
 Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL

- Tissue Dilator: 9 Fr. (3.0 mm) x 10.2 cm

 Dust Cap: Non-Vented

 SecondSite™ Adjustable Hub: Fastener

 SecondSite™ Adjustable Hub: Catheter Clamp
- All components are CE 2797 unless otherwise noted.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

California Prop. 65

▲ WARNING: Cancer and Reproductive Harm

www.P65Warnings.ca.gov

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

Fluid path components are non-pyrogenic

Contraindications: The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.





| Lumen | Priming Volume* (mL) | Gravity Flow Rate† (mL/hr) | |
|-------------------|----------------------------|-------------------------------------|--|
| Distal (14 Ga.) | 0.8 | 5237 | |
| Proximal (14 Ga.) | 0.8 | 3511 | |

Priming volumes are approximate and are done without accessories.



Teleflex Medical EC REP IDA Business and E 2797 Technology Park Dublin Road, Athlone Co. Westmeath, Ireland

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL056518 R01 (2024-02)









(01)XXXXXXXXXXXXXXX (11)YYMMDD (10)LotNumber



20 cm catheter length

.032 inch dia. spring-wire guide







[†] Flow rate values are approximate and are determined using deionized water at 100 cm head height.