



REF PR-05052-LW

16 Ga. 2 Lumen 50 cm catheter length .018 inch dia. spring-wire guide

## Two-Lumen PICC with 130 cm Wire

### Contents:

- 1: Two Lumen Catheter: 16 Ga. (1.76 mm OD) x 50 cm, Blue FlexTip®
  - 1: GlideThru™ Peel-Away Sheath: 5 Fr. x 4" (10 cm) Radiopaque over 5 Fr. Dilator
  - 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on Both Ends)
  - 1: Spring-Wire Guide, Marked: .018" (0.46 mm) dia. x 51-3/16" (130 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other)
  - 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
  - 1: Syringe: 5 mL Luer-Slip
  - 2: Dust Cap: Non-Vented
  - 1: SecondSite™ Adjustable Hub: Fastener
  - 1: SecondSite™ Adjustable Hub: Catheter Clamp
  - 1: Patient ID Card
  - 1: Chart Sticker
  - 1: Patient Information Booklet
  - 1: Paper Tape Measure
  - 1: Dressing: STATLOCK®<sup>1</sup> PICC Plus Catheter Stabilization Device
- <sup>1</sup>A registered trademark of C. R. Bard, Inc.

### Rx only

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com/IFU](http://www.teleflex.com/IFU) California Prop. 65**

**⚠ WARNING:** Cancer and Reproductive Harm.  
[www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

**Not made with natural rubber latex.**

Fluid path components are non-pyrogenic.

**Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.**



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)
Distal (18 Ga.)	0.41	436	3182
Proximal (20 Ga.)	0.39	172	1874

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

LOT



Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

**Teleflex®**  
[www.teleflex.com](http://www.teleflex.com)

LBL056602 R01 (2023-02)



## Two-Lumen PICC with 130 cm Wire

REF PR-05052-LW

16 Ga. 2 Lumen 50 cm catheter length .018 inch dia. spring-wire guide