

Dimensions:  
Width: 8-3/4" (22,2 cm)  
Length: 11-1/8" (28,3 cm)

PB-10001-000A  
PBZ-10001-000A

T-12703-100C  
Width: 8.7598 in = 222.5mm  
Length: 10.374 in = 263.5mm

ARROW



REF PR-05052-LW

(IPNXXXXXX)

16 Ga. 2 Lumen 50 cm catheter length .018 inch dia. spring-wire guide

## Two-Lumen PICC with 130 cm Wire

### Contents:

- 1: Two Lumen Catheter: 16 Ga. (1.76 mm OD) x 50 cm, Blue FlexTip® origin:xxxxxxxxxxxxxxxxxxxx
- 1: GlideThru™ Peel-Away Sheath: 5 Fr. x 4" (10 cm) Radiopaque over 5 Fr. Dilator origin:xxxxxxxxxxxxxxxxxxxx
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on Both Ends) origin:xxxxxxxxxxxxxxxxxxxx
- 1: Spring-Wire Guide, Marked: .018" (0.46 mm) dia. x 51-3/16" (130 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) origin:xxxxxxxxxxxxxxxxxxxx
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW origin:xxxxxxxxxxxxxxxxxxxx
- 1: Syringe: 5 mL Luer-Slip origin:xxxxxxxxxxxxxxxxxxxx
- 2: Dust Cap: Non-Vented origin:xxxxxxxxxxxxxxxxxxxx
- 1: SecondSite™ Adjustable Hub: Fastener origin:xxxxxxxxxxxxxxxxxxxx
- 1: SecondSite™ Adjustable Hub: Catheter Clamp origin:xxxxxxxxxxxxxxxxxxxx
- 1: Patient ID Card origin:xxxxxxxxxxxxxxxxxxxx
- 1: Chart Sticker origin:xxxxxxxxxxxxxxxxxxxx
- 1: Patient Information Booklet origin:xxxxxxxxxxxxxxxxxxxx
- 1: Paper Tape Measure origin:xxxxxxxxxxxxxxxxxxxx
- 1: Dressing: STATLOCK® PICC Plus Catheter Stabilization Device origin:xxxxxxxxxxxxxxxxxxxx

\*A registered trademark of C. R. Bard, Inc.

### Rx only

**Warning:** Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

[www.teleflex.com/IFU](http://www.teleflex.com/IFU)

California Prop. 65

⚠ **WARNING:** Cancer and Reproductive Harm.

[www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

**Not made with natural rubber latex.**

Fluid path components are non-pyrogenic.

**Contraindications:** The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.



Lumen

Priming  
Volume\*  
(mL)

Gravity  
Flow  
Rate†  
(mL/hr)

Pump  
Flow Rate††  
(mL/hr)

Distal (18 Ga.)	0.41	436	3182
Proximal (20 Ga.)	0.39	172	1874

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

LOT

Lot Number



YYYY-MM-DD



YYYY-MM-DD



(01)XXXXXXXXXXXX  
(17)YYMMDD  
(11)YYMMDD  
(10)Lot Number

Set Product of xxx  
Packaged in xxx

Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

Teleflex®  
[www.teleflex.com](http://www.teleflex.com)

LBL056602 R02 (2024-05)



## Two-Lumen PICC with 130 cm Wire

ARROW

REF PR-05052-LW

16 Ga. 2 Lumen

50 cm catheter length

.018 inch dia. spring-wire guide