

Dimensions:
 Width: 8-3/4" (22,2 cm)
 Length: 11-1/8" (28,3 cm)

PB-10001-000A
PBZ-10001-000A

T-12703-100C
 Width: 8.7598 in = 222.5mm
 Length: 10.374 in = 263.5mm



REF PR-05041-LW (IPNXXXXXX)

4 Fr. **1** Lumen **50** cm catheter length **.018** inch dia. spring-wire guide

One-Lumen PICC with 130 cm Wire

Contents:

- 1: One Lumen Catheter: 4 Fr. (1.45 mm OD) x 50 cm, Blue FlexTip® origin:xxxxxxxxxxxxxxxxxxxx
- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 4" (10 cm) Radiopaque over 4 Fr. Dilator origin:xxxxxxxxxxxxxxxxxxxx
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on Both Ends) origin:xxxxxxxxxxxxxxxxxxxx
- 1: Spring-Wire Guide, Marked: .018" (0.46 mm) dia. x 51-3/16" (130 cm)(Straight Soft Tip on One End - Straight Stiff Tip on Other) origin:xxxxxxxxxxxxxxxxxxxx
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW origin:xxxxxxxxxxxxxxxxxxxx
- 1: Syringe: 5 mL Luer-Slip origin:xxxxxxxxxxxxxxxxxxxx
- 1: Dust Cap: Non-Vented origin:xxxxxxxxxxxxxxxxxxxx
- 1: SecondSite™ Adjustable Hub: Catheter Clamp origin:xxxxxxxxxxxxxxxxxxxx
- 1: SecondSite™ Adjustable Hub: Fastener origin:xxxxxxxxxxxxxxxxxxxx
- 1: Patient ID Card origin:xxxxxxxxxxxxxxxxxxxx
- 1: Chart Sticker origin:xxxxxxxxxxxxxxxxxxxx
- 1: Patient Information Booklet origin:xxxxxxxxxxxxxxxxxxxx
- 1: Paper Tape Measure origin:xxxxxxxxxxxxxxxxxxxx
- 1: Dressing: STATLOCK®¹ PICC Plus Catheter Stabilization Device origin:xxxxxxxxxxxxxxxxxxxx

¹A registered trademark of C. R. Bard, Inc.

Rx only

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

www.teleflex.com/IFU

California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm.
www.P65Warnings.ca.gov

Not made with natural rubber latex.

Fluid path components are non-pyrogenic.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)
Distal (18 Ga.)	0.50	1029	4708

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

Arrow International LLC
 Subsidiary of Teleflex Incorporated
 3015 Carrington Mill Blvd.
 Morrisville, NC 27560 USA

LOT Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXX
 (17)YYMMDD
 (11)YYMMDD
 (10)Lot Number

Set Product of xxx
 Packaged in xxx

Teleflex
www.teleflex.com

LBL056613 R02 (2024-05)



One-Lumen PICC with 130 cm Wire



REF PR-05041-LW

4 Fr. **1** Lumen **50** cm catheter length **.018** inch dia. spring-wire guide