

Product of xxx  
Packaged in xxx



REF **PR-05042** (IPNXXXXXX)

**4** Fr. **2** Lumen **50** cm catheter length **.018** inch dia. spring-wire guide

# Two-Lumen PICC

**Contents:**

- 1: Two Lumen Catheter: 4 Fr. (1.45 mm OD) x 50 cm, Blue FlexTip®
- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 4" (10 cm) Radiopaque over 4 Fr. Dilator
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other)
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 5 mL Luer-Slip
- 2: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: Chart Sticker
- 1: Patient Information Booklet
- 1: Patient ID Card
- 1: Paper Tape Measure
- 1: Dressing: STATLOCK®<sup>1</sup> Catheter Stabilization Device

**Rx only**

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU California Prop. 65**

**⚠ WARNING: Cancer and Reproductive Harm. www.P65Warnings.ca.gov**

**Not made with natural rubber latex.**

Fluid path components are non-pyrogenic.

**Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.**

<sup>1</sup>A registered trademark of C. R. Bard, Inc.

**All components are CE 2797 unless otherwise noted.**



CAS # 7440-48-4

en es

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)
Distal (22 Ga.)	0.27	165	1446
Proximal (22 Ga.)	0.33	139	1529

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

EU Authorized Representative and Importer:

Teleflex Medical  
IDA Business and Technology Park  
Dublin Road, Athlone  
Co. Westmeath, Ireland

Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

**Teleflex®**  
www.teleflex.com

LBL056842 R02 (2024-01)

LOT Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXXX  
(17)YYMMDD  
(11)YYMMDD  
(10)Lot Number



**4** Fr. **2** Lumen **50** cm catheter length **.018** inch dia. spring-wire guide

REF PR-05042

# Two-Lumen PICC

