



REF CDC-15123-P1A

12 Fr. 3 Lumen 20 cm catheter length .035 inch dia. spring-wire guide

ARROW

ErgoPack®

Pressure Injectable Three-Lumen Hemodialysis Catheterization for High Volume Infusions

Contents:

- | | |
|--|---|
| <ul style="list-style-type: none"> 1: Three-Lumen Catheter: 12 Fr. (4.06 mm OD) x 20 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle 1: Injection Needle: SafetyGlide™¹ 25 Ga. x 1" (2.54 cm) 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) 1: Pressure Transduction Probe 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe 1: Injection Needle: SafetyGlide™¹ 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe 1: Syringe: 3 mL Luer-Lock 1: Tissue Dilator: 10 Fr. (3.5 mm) x 10.2 cm 1: Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®² One-Step Solution with Hi-Lite Orange™ Tint 1: SharpsAway® Disposal Cup 1: SharpsAway® II Locking Disposal Cup | <ul style="list-style-type: none"> 1: Maximal Barrier Drape™ with 4" fenestration 1: Needle Holder 1: Safety Scalpel: #11 1: Checklist/CLIP Sheet 1: Sterile Procedure Sign 1: Medication Label: 1% Lidocaine 2: Gauze Pad: 2" x 2" (5 cm x 5 cm) 5: Gauze Pad: 4" x 4" (10 cm x 10 cm) 1: Surgical Apparel: Impervious Gown 1: Surgical Apparel: Bouffant Cap 1: Surgical Apparel: Mask with Eye Shield 1: Suture: 3-0 Silk with Curved Needle 1: HemoHopper® Fluid Receptacle <p>¹A trademark of Becton, Dickinson and Company.
 ²A registered trademark of CareFusion or one of its subsidiaries.</p> |
|--|---|

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Contraindications: The Multi-Lumen Acute Hemodialysis Catheter (AHDC) is not designed for long-term (≥ 30 days) hemodialysis or for insertion into thrombosed vessels.

