PB-10000-850 8.5 X 10.5



REF CDC-26142-1A (IPNXXXXXX)

Arrowg+ard Blue[®] Two-Lumen Hemodialysis Catheterization for High Volume Infusions

25^{cm}_{catheter}

Contents:

- Two-Lumen Catheter: 14 Fr. (4.72 mm OD) x 25 cm Radiopaque Polyurethane with Blue FlexTip®, Arrowg+ard® Antimicrobial Surface Treatment, Extension Line Clamps Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight
- Soft Tip on One End "J" Tip on Other) with Arrow Advancer Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer
- Needle
- Injection Needle: SafetyGlide^{™1} 25 Ga. x 1" (2.54 cm) Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) 1:
- Pressure Transduction Probe
- Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- Injection Needle: SafetyGlide™1 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Svrinae
- Syringe: 3 mL Luer-Lock
- Tissue Dilator: 14 Fr. (4.7 mm) x 14 cm 1:
- Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
- 5 mL 1% Lidocaine HCI Solution and Alcohol Prep 1.
- 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®2 One-Step Solution with Hi-Lite Orange™ Tint
- SharpsAway® II Locking Disposal Cup

1: SharpsAway® Disposal Cup

Maximal Barrier Drape™ with 4" fenestration 1.

.035 inch dia. spring-wire quide

- Needle Holder
- 1: Safety Scalpel: #11 1:
- Checklist/CLIP Sheet 1:
- Sterile Procedure Sign 1:
- Medication Label: 1% Lidocaine 1.
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5. Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Surgical Apparel: Impervious Gown
- Surgical Apparel: Bouffant Cap 1:
- Surgical Apparel: Mask with Eye Shield 1:
- Suture: 3-0 Braided Silk with Straight Needle 1:
- Suture: 3-0 Silk with Curved Needle 1.
- 1: HemoHopper® Fluid Receptacle
- ¹A trademark of Becton, Dickinson and Company ²A registered trademark of CareFusion or one of its subsidiaries.

Not made with natural rubber latex.


Store between 20 - 25°C (68 - 77°F)

20°C (77°F)

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term (≥ 30 days) hemodialysis or for use in patients with thrombosed vessels

The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/ or sulfa drugs.







LBL057046 R01 (2024-05)

ARROW

Erg[®]Pack