



REF CDC-12123-P1A

12 Fr.

3 Lumen

16 cm catheter length

.035 inch dia. spring-wire guide

ARROW

ErgoPack

Pressure Injectable Three-Lumen Hemodialysis Catheterization for High Volume Infusions

Contents:

- | | |
|--|--|
| 1: Three-Lumen Catheter: 12 Fr. (4.06 mm OD) x 16 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps | 1: Maximal Barrier Drape™ with 4" fenestration |
| 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer | 1: Needle Holder |
| 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle | 1: Safety Scalpel: #11 |
| 1: Injection Needle: SafetyGlide™ 1 25 Ga. x 1" (2.54 cm) | 1: Checklist/CLIP Sheet |
| 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) | 1: Sterile Procedure Sign |
| 1: Pressure Transduction Probe | 1: Medication Label: 1% Lidocaine |
| 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe | 2: Gauze Pad: 2" x 2" (5 cm x 5 cm) |
| 1: Injection Needle: SafetyGlide™ 1 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe | 5: Gauze Pad: 4" x 4" (10 cm x 10 cm) |
| 1: Syringe: 3 mL Luer-Lock | 1: Surgical Apparel: Impervious Gown |
| 1: Tissue Dilator: 10 Fr. (3.5 mm) x 10.2 cm | 1: Surgical Apparel: Bouffant Cap |
| 1: Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm | 1: Surgical Apparel: Mask with Eye Shield |
| 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep | 1: Suture: 3-0 Silk with Curved Needle |
| 1: 3 mL Applicator 2% CHG and 70% IPA ChlorPrep® ² One-Step Solution with Hi-Lite Orange™ Tint | 1: HemoHopper® Fluid Receptacle |
| 1: SharpsAway® Disposal Cup | |
| 1: SharpsAway® II Locking Disposal Cup | |

¹A trademark of Becton, Dickinson and Company.
²A registered trademark of CareFusion or one of its subsidiaries.

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Contraindications: The Multi-Lumen Acute Hemodialysis Catheter (AHDC) is not designed for long-term (≥ 30 days) hemodialysis or for insertion into thrombosed vessels.

