





(IPNXXXXXX)



Pressure Injectable Three-Lumen Hemodialysis Catheterization for High Volume **Infusions**



- Three-Lumen Catheter: 12 Fr. (4.06 mm OD) x 25 cm Pressure Injectable
- Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer
- Injection Needle: SafetyGlide™1 25 Ga. x 1" (2.54 cm)
- Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- Pressure Transduction Probe
- Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- Injection Needle: SafetyGlide™1 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- Syringe: 3 mL Luer-Lock
- Tissue Dilator: 10 Fr. (3.5 mm) x 10.2 cm
- Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
- 10 mL Luer-Lock Pre-Filled Saline Syringe
- 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®2 One-Step Solution with Hi-Lite Orange™ Tint
- SharpsAway® II Locking Disposal Cup

- SharpsAway® Disposal Cup
- 1: Maximal Barrier Drape™ with 4" fenestration
- 1: Needle Holder
- 1: Safety Scalpel: #11
- Checklist/CLIP Sheet
- Sterile Procedure Sign
- 1: Medication Label: 1% Lidocaine
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- Gauze Pad: 4" x 4" (10 cm x 10 cm) Dressing: Tegaderm®3 10 cm x 12 cm
- Surgical Apparel: Impervious Gown
- Dressing: BIOPATCH®4
- Surgical Apparel: Bouffant Cap
- Surgical Apparel: Mask with Eye Shield
- Suture: 3-0 Silk with Curved Needle
- 1: HemoHopper® Fluid Receptacle
- ¹A trademark of Becton, Dickinson and Company
- ²A registered trademark of CareFusion or one of its subsidiaries.
- ³A registered trademark of 3M Company.
- ⁴A registered trademark of Johnson & Johnson Corporation.

Not made with natural rubber latex.

Store between 20 - 25°C (68 - 77°F)

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and

Contraindications: The Multi-Lumen Acute Hemodialysis Catheter (AHDC) is not designed for long-term (≥ 30 days) hemodialysis or for insertion into thrombosed vessels



LBL057052 R01 (2024-06)