



REF CDC-22142-1A

(IPNXXXXXX)

14^{Fr.}

2^{Lumen}

15^{cm}
catheter
length

.035^{inch dia.}
spring-wire
guide

ARROW®

ErgoPack®



Arrowg+ard Blue® Two-Lumen Hemodialysis Catheterization for High Volume Infusions

Contents:

- 1: Two-Lumen Catheter: 14 Fr. (4.72 mm OD) x 15 cm Radiopaque Polyurethane with Blue FlexTip®, Arrowg+ard® Antimicrobial Surface Treatment, Extension Line Clamps
- 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- 1: Injection Needle: SafetyGlide™¹ 25 Ga. x 1" (2.54 cm)
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Pressure Transduction Probe
- 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- 1: Injection Needle: SafetyGlide™¹ 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- 1: Syringe: 3 mL Luer-Lock
- 1: Tissue Dilator: 14 Fr. (4.7 mm) x 14 cm
- 1: Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
- 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 1: 3 mL Applicator 2% CHG and 70% IPA Chloraprep®² One-Step Solution with Hi-Lite Orange™ Tint
- 1: SharpsAway® II Locking Disposal Cup

- 1: SharpsAway® Disposal Cup
- 1: Maximal Barrier Drape™ with 4" fenestration
- 1: Needle Holder
- 1: Safety Scalpel: #11
- 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign
- 1: Medication Label: 1% Lidocaine
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Surgical Apparel: Impervious Gown
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Suture: 3-0 Braided Silk with Straight Needle
- 1: Suture: 3-0 Silk with Curved Needle
- 1: HemoHopper® Fluid Receptacle
- ¹A trademark of Becton, Dickinson and Company.
- ²A registered trademark of CareFusion or one of its subsidiaries.


Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term (≥ 30 days) hemodialysis or for use in patients with thrombosed vessels.

The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/or sulfa drugs.




Priming Volume*

Distal	1.40 mL
Proximal	1.10 mL

*Priming volumes are approximate and are done without accessories.

Dialysis Flow Rate†



† In vitro tests performed using 3:1 glycerine/saline at 37°C.

