



REF ASK-46122-VG



12 Fr. 2 Lumen 25 cm catheter length .035 inch dia. spring-wire guide

Arrowg+ard Blue® Two-Lumen Hemodialysis Catheterization Kit for High Volume Infusions

Contents:

- 1: Two-Lumen Catheter: 12 Fr. (4.06 mm OD) x 25 cm Radiopaque Polyurethane with Blue FlexTip®, Arrowg+ard® Antimicrobial Surface Treatment, Extension Line Clamps
- 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- 1: Pressure Transduction Probe
- 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- 1: Injection Needle: Eclipse™ 1 23 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- 1: Injection Needle: Eclipse™ 1 25 Ga. x 1" (2.54 cm) and 3 mL Luer-Lock Syringe
- 1: Syringe: 10 mL Luer-Lock
- 1: Tissue Dilator: 10 Fr. (3.5 mm) x 10.2 cm
- 1: Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
- 1: SharpsAway® II Locking Disposal Cup
- 1: SharpsAway® Disposal Cup
- 1: Maximal Barrier Drape™ with 4" fenestration
- 1: Towel
- 1: Needle Holder
- 1: Forceps
- 1: Safety Scalpel: #11
- 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm® 10 cm x 12 cm
- 1: Surgical Apparel: Impervious Gown
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Suture: 3-0 Braided Silk with Straight Needle
- 1: Suture: 3-0 Silk with Curved Needle
- 1: HemoHopper® Fluid Receptacle


¹A trademark of Becton, Dickinson and Company.
²A registered trademark of 3M Company.

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).


Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term (≥ 30 days) hemodialysis or for use in patients with thrombosed vessels.

The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/or sulfa drugs.



Priming Volume*	
Distal	1.40 mL
Proximal	1.60 mL

*Priming volumes are approximate and are done without accessories.



† In vitro tests performed using 3:1 glycerine/saline at 37°C.

