





(IPNXXXXXX)

16 cm catheter length .035 inch dia. spring-wire quide

## Arrowg+ard Blue® You-Bend™ Two-Lumen Hemodialysis Catheterization for High **Volume Infusions**

- Two-Lumen Catheter‡: 12 Fr. (4.06 mm OD) x 16 cm Radiopaque Polyurethane with Blue FlexTip®

- origin:xxxxxxxxxxxxxxxxxxxxxxxx

- <sup>1</sup>A registered trademark of CareFusion or one of its subsidiaries.

Rx only					
Varning: Read all	l package insert wa	rnings, precaution	s, and instructions	prior to use.	Failure

e to do so may result in severe patient injury or death.

California Prop. 65

A WARNING: Cancer and Reproductive Harm.

www.P65Warnings.ca.gov ‡MR Conditional

Not made with natural rubber latex.

Store between  $20 - 25^{\circ}\text{C}$  (68 - 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term (≥ 30 days) hemodialysis or for use in patients with thrombosed vessels.

Lumen	Priming Volume* (mL)	Flow Rate (mL/min)	Venous Pressure Setting†† (mmHg)
Distal (12 Ga.)	1.6	220	100
		370	200
Proximal (12 Ga.)	1.5	220	100
,		370	200

\* Priming volumes are approximate and are done without accessories †† Blood flow rates are approximate and were determined by varying the venous pressure setting from 100 to 200 mmHg in an in vitro experiment using blood analog solution.

The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/or sulfa drugs.



M YYYY-MM-DD

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA

Set Product of xxx Packaged in xxx

**Teleflex**® LBL057088 R01 (2024-05)



YYYY-MM-DD







.035 inch dia. spring-wire guide





(11)YYMMDD

