











## Pressure Injectable Arrowg+ard Blue Advance® One-Lumen Midline

## Contents:

- TaperFree® Catheter with Arrowg+ard Blue Advance™ Antimicrobial/Antithrombogenic Protection¹: 4.5 Fr. (1.6 mm
- OD) x 15 cm, Pressure Injectable, Blue FlexTip®
  GlideThru™ Peel-Away Sheath: 4.5 Fr. x 2-3/4" (7 cm) Radiopaque over 4.5 Fr. Dilator
  Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End Straight
- Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- Syringe: 10 mL Luer-Lock
- Syling State Cook
  Dust Cap: Non-Vented
  SecondSite™ Adjustable Hub: Catheter Clamp
  SecondSite™ Adjustable Hub: Fastener
- SharpsAway® II Locking Disposal Cup SharpsAway® Disposal Cup
- Catheter Trimmer
- Safety Scalpel: #11
- Patient ID Card
- Patient Information Booklet
- Paper Tape Measure
- Dressing: STATLOCK®2 PICC Plus Catheter Stabilization Device

<sup>1</sup>Licensed under US Patent No. 7,329,412.

<sup>2</sup>A registered trademark of C. R. Bard, Inc.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm.

www.P65Warnings.ca.gov

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).

Fluid path components are non-pyrogenic.

Contraindications: The Arrowg+ard Blue Advance Pressure Injectable Midline is contraindicated for patients with known hypersensitivity to chlorhexidine.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (17 Ga.)	0.31	2009	8820	5

<sup>\*</sup> Priming volumes are approximate and are done without accessories.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.


\*\* Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with 152 cm administration set tubing.




Morrisville, NC 27560 USA Product of xxx Packaged in xxx

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd.









(17)YYMMDD (11)YYMMDD (10)Lot Number

LBL057532 R01 (2024-11)

<sup>†</sup> Flow rate values are approximate and are determined using deionized water at 100 cm head height.