



## REF MP-14702-JA

30 cm catheter length

.018 spring-wire guide

## Two-Lumen CVC

## Contents:

- Two-Lumen Catheter: 7 Fr. (2.5 mm OD) x 30 cm,
- Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 27-1/2" (70 cm) (Angled Soft Tip on One End -Straight Stiff Tip on Other) with Arrow Advancer and
- Catheter: 20 Ga. x 1-3/4" (4.45 cm) Radiopaque over 22 Ga. TW Introducer Needle
- Injection Needle: 25 Ga. x 1-1/2" (3.81 cm) Injection Needle: 22 Ga. x 1-1/2" (3.81 cm)
- Introducer Needle: Echogenic 22 Ga. (0.72 mm) x 2-3/4" (7 cm) XTW
  - Introducer Needle: Echogenic 22 Ga. (0.72 mm) x 1-3/8" (3.5 cm) XTW
- Syringe: 10 mL Luer-Slip
- Syringe: 5 mL Luer-Slip
- 1: Syringe: 2 mL Luer-Slip
- Tissue Dilator: 8 Fr. x 9.7 cm Dust Cap: Non-Vented
- SecondSite™ Adjustable Hub: Fastener

Fluid path components are non-pyrogenic.

- SecondSite™ Adjustable Hub: Catheter Clamp 1:
- Medicine Cup: 60 mL

SharpsAway® Disposal Cup

- Drape: 35" x 50" (89 cm x 127 cm) with 4" (10 cm) fenestration, with adhesive
- Needle Holder
- Scalpel: #11
- 2: Stopcock: 4-way

- Prep Sponge Swab
  Gauze Pad: 4" x 4" (10 cm x 10 cm)
  Dressing: OpSite®1 4-3/4" x 4" (12 cm x 10 cm) 1:
- Suture: 3-0 Silk with Curved Needle
- Tray: Prep
- Y-hub
- Guidewire Introducer

<sup>1</sup>A registered trademark of T.J. Smith & Nephew, Limited.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.  Not made with natural rubber latex.  Sterilized using ethylene oxide unless otherwise indicated in the contents	6 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
list.	Distal (16 Ga.) Proximal (16 Ga.)	0.57 0.62	1917 2000	

<sup>\*</sup> Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.





Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL057809 R00 (2021-10)



