



REF MPH-15703-JA

7 Fr. 3 Lumen 20 cm catheter length .018 inch dia. spring-wire guide

## Pressure Injectable Three-Lumen CVC

**Contents:**

- 1: Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 20 cm, Pressure Injectable, Stylet
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 23-5/8" (60 cm) (Angled Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer and hub
- 1: Catheter: 20 Ga. x 1-3/4" (4.45 cm) Radiopaque over 22 Ga. TW Introducer Needle
- 1: Injection Needle: 25 Ga. x 1-1/2" (3.81 cm)
- 1: Injection Needle: 22 Ga. x 1-1/2" (3.81 cm)
- 1: Introducer Needle: Echogenic 22 Ga. (0.72 mm) x 2-3/4" (7 cm) XTW
- 1: Introducer Needle: Echogenic 22 Ga. (0.72 mm) x 1-3/8" (3.5 cm) XTW
- 1: Syringe: 10 mL Luer-Slip
- 1: Syringe: 5 mL Luer-Slip
- 1: Syringe: 2 mL Luer-Slip
- 1: Tissue Dilator: 8 Fr. x 9.7 cm
- 3: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: Medicine Cup: 60 mL
- 1: SharpsAway® Disposal Cup
- 1: Drape: 35" x 50" (89 cm x 127 cm) with 4" (10 cm) fenestration, with adhesive
- 1: Needle Holder
- 1: Scalpel: #11
- 2: Stopcock: 4-way
- 1: Prep Sponge Swab
- 6: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: OpSite® 4-3/4" x 4" (12 cm x 10 cm)
- 1: Suture: 3-0 Silk with Curved Needle
- 1: Tray: Prep
- 1: Y-hub
- 2: Guidewire Introducer

<sup>1</sup>A registered trademark of T.J. Smith & Nephew, Limited.

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.**

**Not made with natural rubber latex.**

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.



Lumen

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (16 Ga.)	0.47	2656	10
Medial (18 Ga.)	0.4	1245	5
Proximal (18 Ga.)	0.42	1328	5

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

\*\* Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with 152 cm administration set tubing.



Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

**Teleflex®**  
www.teleflex.com

LBL057815 R00 (2021-10)

