





REF CS-26122-F (IPNXXXXXXX)

25 cm catheter length .035 inch dia. spring-wire guide

Arrowg+ard Blue® Two-Lumen Hemodialysis Catheter for High **Volume Infusions**

- Two-Lumen Catheter: 12 Fr. (4.06 mm OD) x 25 cm Radiopaque Polyurethane with Blue FlexTip®, Arrowg+ard® Antimicrobial Surface Treatment, Extension Line Clamps
- Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer
 Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
 Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW
 Pressure Transduction Probe

- Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL
- Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm Dust Cap: Non-Vented Drape: 24" x 36" (61 cm x 91 cm)

- Scalpel: #11

All components are CE 2797 unless otherwise noted

Rx	on	ŀ

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm.

www.P65Warnings.ca.gov

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

Fluid path components are non-pyrogenic.

Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term (≥ 30 days) hemodialysis or for use in

patients with thrombosed vessels.

The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/or sulfa drugs.

Lumen	Priming Volume* (mL)	Flow Rate (mL/min)	Venous Pressure Setting†† (mmHg)
Distal (12 Ga.)	1.8	170	100
		300	200
Proximal (12 Ga.)	1.7	170	100
		300	200

^{*} Priming volumes are approximate and are done without accessories.

^{††} Blood flow rates are approximate and were determined by varying the venous pressure setting from 100 to 200 mmHg in an in vitro experiment using blood analog pressure solution.




































Technology Park
Dublin Road, Athlone
Co. Westmeath, Ireland All components are CE 2797 unless otherwise noted.


Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



ARROW

LOT Lot Number







(01)XXXXXXXXXXXXXXX (17)YYMMDD (11)YYMMDD (10)Lot Number


Product of xxx Packaged in xxx











25 cm catheter length

.035 inch dia. spring-wire guide