





(IPNXXXXXX)

9 Fr. distal lumen



10 cm device length

■035 inch dia. spring-wire guide

MAC™ Two-Lumen Central Venous Access for use with 7 - 7.5 Fr. Catheters

Contents:

- origin:xxxxxxxxxxxxxxxxxxxxxx
- Cath-Gard® Catheter Contamination Shield: 80 cm with Tuohy-Borst Adapter, Sterile Tape origin:xxxxxxxxxxxxxxxxxxxxxx

- ¹A registered trademark of CareFusion or one of its subsidiaries.

Rx only
Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU
California Prop. 65

\(\triangle \) WARNING: Cancer and Reproductive Harm.
\(\text{www.P65Warnings.ca.gov} \)

\(\text{Precaution: Contains Phthalate: DEHP.} \)
Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Store between 20 - 25°C (68 - 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents

Fluid path components are non-pyrogenic.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (9 Fr.) no catheter	1.65	31440
Distal (9 Fr.) with 8 Fr. catheter	0.9	9040
Proximal (12 Ga) no catheter	0.6	8970

^{*} Priming volumes are approximate and are done without accessories.



Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA

Teleflex®

LBL058364 R01 (2024-04)



LOT Lot Number

YYYY-MM-DD



YYYY-MM-DD





10 cm device length

.035 inch dia. spring-wire quide





(01)XXXXXXXXXXXXXXX (17)YYMMDD (11)YYMMDD

(10)Lot Number





[†] Flow rate values are approximate and are determined using deionized water at 100 cm head height.