



REF NA-04020-S1A

20^{Ga.} 4.45^{cm} catheter length .018^{inch dia.} spring-wire guide

Radial Artery Catheterization

Contents:

- | | |
|--|---|
| <ul style="list-style-type: none"> 1: Catheter: 20 Ga. x 4.45 cm Radiopaque Polyurethane over 22 Ga. TW 1: Introducer Needle with Integral .018" (0.46 mm) dia. Spring-Wire Guide 1: Injection Needle: SafetyGlide™¹ 25 Ga. x 1" (2.54 cm) 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) 1: Syringe: 3 mL Luer-Lock 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®² One-Step Solution with Hi-Lite Orange™ Tint 1: SharpsAway® Disposal Cup 1: SharpsAway® II Locking Disposal Cup 1: Drape: 36" x 41" (91 cm x 104 cm) with 3" x 5" (8 cm x 13 cm) fenestration 1: Towel: 13" x 18" (33 cm x 46 cm) 1: Checklist Sheet 1: Medication Label: 1% Lidocaine 1: Dressing: GRIP-LOK®³ Catheter Fixation Device 4: Gauze Pad: 2" x 2" (5 cm x 5 cm) 4: Gauze Pad: 4" x 4" (10 cm x 10 cm) 1: Dressing: I.V. Advanced 10 cm x 12 cm 1: Dressing: BIOPATCH®⁴ 1: Surgical Apparel: Mask with Eye Shield | <ul style="list-style-type: none"> 1: Surgical Apparel: Bouffant Cap 1: Tubing±: 6" Extension with One-Way Stopcock <p>¹A trademark of Becton, Dickinson and Company.
 ²A registered trademark of CareFusion or one of its subsidiaries.
 ³A registered trademark of TIDI.
 ⁴A registered trademark of Johnson & Johnson Corporation.</p> |
|--|---|

Rx only

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm.
www.P65Warnings.ca.gov

±Precaution: Contains Phthalate: DEHP.

Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.



LOT



Arrow International LLC
 Subsidiary of Teleflex Incorporated
 3015 Carrington Mill Blvd.
 Morrisville, NC 27560 USA

Teleflex®
www.teleflex.com

LBL059314 R01 (2022-02)



REF NA-04020-S1A

20^{Ga.} 4.45^{cm} catheter length .018^{inch dia.} spring-wire guide

Radial Artery Catheterization