

Dimensions:
Width: 10-5/8"
Length: 12-1/2"

PB-10003-000

T04400101A3 Dimensions: 10 1/2" x 10.5"

K08800109A3 Dimensions: 10 1/2" x 11"



REF MTO-11142-AJ (IPNXXXXXX)




9 Fr. distal lumen 2 Lumen 10 cm device length .035 inch dia. spring-wire guide

MAC™ Two-Lumen Central Venous Access for use with 7.5 - 8 Fr. Catheters

Contents:

- 1: Two-Lumen Central Venous Access Device: 4" (10 cm) Radiopaque Polyurethane with 9 Fr. Distal Lumen, Integral Hemostasis Valve, Extension Line Clamps
1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
1: Injection Needle: 22 Ga. x 1-1/2" (3.81 cm)
1: Injection Needle: 25 Ga. x 1" (2.54 cm)
1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW
1: Pressure Transduction Probe
1: Syringe: 3 mL Luer-Lock
1: Syringe: 5 mL Luer-Slip
1: Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL
1: Tissue Dilator: 9 Fr. (3.0 mm) x 20.7 cm
1: Obturator: 8 Fr.
1: Cath-Gard® Catheter Contamination Shield: 80 cm with TwistLock™ Adapter
1: Dust Cap: Vented
2: Dust Cap: Non-Vented
1: SharpsAway® Disposal Cup
1: Drape: 24" x 36" (61 cm x 91 cm)
1: Drape: 21" x 36" (53 cm x 91 cm) with 3" (7.6 cm) fenestration, with adhesive
1: Scalpel: #11
1: Stopcock: 4-way High-Flow
2: Prep Sponge Swab
2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
1: Suture: 3-0 Braided Silk with Straight Needle
1: Tray: Prep

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.
Not made with natural rubber latex.
Sterilized using ethylene oxide unless otherwise indicated in the contents list.
Fluid path components are non-pyrogenic.

 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (9 Fr.) no catheter	1.65	31440
Distal (9 Fr.) with 8 Fr. catheter	0.9	9040
Proximal (12 Ga.) no catheter	0.6	8970

* Priming volumes are approximate and are done without accessories.
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

LOT Lot Number

YYYY-MM-DD

YYYY-MM-DD

(01)XXXXXXXXXXXXX
(17)YYMMDD
(11)YYMMDD
(10)Lot Number

Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Product of xxx
Packaged in xxx

Teleflex®
www.teleflex.com
LBL059318 R02 (2024-10)



REF MTO-11142-AJ

9 Fr. distal lumen 2 Lumen 10 cm device length .035 inch dia. spring-wire guide

MAC™ Two-Lumen Central Venous Access for use with 7.5 - 8 Fr. Catheters

