



REF NA-04220-X1A (IPNXXXXXX)

20 Ga. 3.81 cm catheter length of the spring-wire guide

QuickFlash® Radial Artery Catheterization

- ¹A trademark of Becton, Dickinson and Company.

²A registered trademark of CareFusion or one of its subsidiaries.
 ³A registered trademark of 3M Company.
 ⁴A registered trademark of Johnson & Johnson Corporation.

Warning: Read all package insert warnings, precautions, and

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU
California Prop. 65

\(\triangle WARNING: Cancer and Reproductive Harm. \)
www.P65Warnings.ca.gov

\(\triangle Precaution: Contains Phthalate: DEHP. \)
Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.

Not made with natural rubber latex.

Store between 20 - 25°C (68 - 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents

Fluid path components are non-pyrogenic.

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.




Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA

Teleflex®

LBL059319 R02 (2024-05)

LOT Lot Number YYYY-MM-DD





(01)XXXXXXXXXXXXXXX (17)YYMMDD (11)YYMMDD (10)Lot Number

