





20 Ga. 4.45 cm catheter enter quide

Radial Artery Catheterization

- Catheter: 20 Ga. x 4.45 cm Radiopaque Polyurethane over 22 Ga. TW Introducer Needle with Integral .018" (0.46 mm) dia. Spring-Wire Guide Injection Needle: SafetyGlide™1 25 Ga. x 1" (2.54 cm) Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip

- Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip Syringe: 3 mL Luer-Lock
 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®2 One-Step Solution with Hi-Lite Orange™ Tint
 Wing Clip: Suture
 SharpsAway® II Locking Disposal Cup
 Towel: 13" x 18" (33 cm x 46 cm)
 Drape: 18" x 26" with 2" x 3" fenestration, with adhesive
 Scalpel: #11
 Gauze Pad: 2" x 2" (5 cm x 5 cm)
 Independent of Becton, Dickinson and Company

- ¹A trademark of Becton, Dickinson and Company. ²A registered trademark of CareFusion or one of its subsidiaries.

Rx only Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/

IFU
California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm.
www.P65Warnings.ca.gov
Not made with natural rubber latex.
Store between 15 – 30°C (59 – 86°F). Avoid freezing and excessive heat above 40°C (104°F).
Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.





20 Ga. 4.45 cm catheter length .018 spring-wire guide

Radial Artery Catheterization

- Catheter: 20 Ga. x 4.45 cm Radiopaque Polyurethane over 22 Ga. TW Introducer Needle with Integral .018" (0.46 mm) dia. Spring-Wire Guide Injection Needle: SafetyGlide™1 25 Ga. x 1" (2.54 cm) Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip

- Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip Syringe: 3 mL Luer-Lock
 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®2 One-Step Solution with Hi-Lite Orange™ Tint
 Wing Clip: Suture
 SharpsAway® II Locking Disposal Cup
 Towel: 13" x 18" (33 cm x 46 cm)
 Drape: 18" x 26" with 2" x 3" fenestration, with adhesive
 Scalpel: #11
 Gauze Pad: 2" x 2" (5 cm x 5 cm)

- ¹A trademark of Becton, Dickinson and Company. ²A registered trademark of CareFusion or one of its subsidiaries.

Rx only Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/ IFU

IFU
California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm.
www.P65Warnings.ca.gov
Not made with natural rubber latex.
Store between 15 – 30°C (59 – 86°F). Avoid freezing and excessive heat above 40°C (104°F).
Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.







Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL059345 R00 (2022-03)

ARROW







Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA



LBL059345 R00 (2022-03)

ARROW













