



## **QuickFlash® Radial Artery Catheterization**

- QuickFlash® Radial Artery Catheter: 20 Ga. x 1-1/2" (3.81 cm) Radiopaque Polyurethane over 21 Ga. TW Needle with Integral .018" (0.46 mm) dia. Spring-Wire Guide
- Injection Needle: SafetyGlide™1 25 Ga. x 1" (2.54 cm)

- Injection Needle: SafetyGlide: "" | 25 Ga. x 1" (2.54 cm)
  Syringe: 3 mL Luer-Lock
  SharpsAway® Disposal Cup
  SharpsAway® II Locking Disposal Cup
  Drape: 36" x 41" (91 cm x 104 cm) with 3" x 5" (8 cm x 13 cm) fenestration
  Towel: 13" x 18" (33 cm x 46 cm)
- Needle Holder

- Needle Holder
  Safety Scalpel: #11
  Medication Label: 1% Lidocaine
  Gauze Pad: 2" x 2" (5 cm x 5 cm)
  Gauze Pad: 4" x 4" (10 cm x 10 cm)
  Dressing: I.V. Advanced 10 cm x 12 cm

- Surgical Apparel: Bouffant Cap
  Surgical Apparel: Mask with Eye Shield
  Suture: 3-0 Silk with Curved Needle
  Tubing±: 6" Extension with One-Way Stopcock
- <sup>1</sup>A trademark of Becton, Dickinson and Company,

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

±Precaution: Contains Phthalate: DEHP.

Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.

Fluid path components are non-pyrogenic.

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.



LOT



M



Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL059372 R00 (2022-02)





