







40 cm catheter length 1018 inch dia. spring-wire quide

Pressure Injectable One-Lumen PICC

Contents:

- One Lumen Catheter: 4 Fr. (1.45 mm OD) x 40 cm, Pressure Injectable, T-
- Port Connector, Blue FlexTip®, Contamination Guard and Placement Wire GlideThru™ Peel-Away Sheath: 4 Fr. x 2-3/4" (7 cm) Radiopaque over 4 Fr. Dilator
- Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End Straight Stiff Tip on Other) with Arrow
- Advancer
 Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
 Syringe: 10 mL Luer-Lock
 SecondSite™ Adjustable Hub: Catheter Clamp
 SecondSite™ Adjustable Hub: Fastener
 SharpsAway® II Locking Disposal Cup

- Catheter Trimmer
- Safety Scalpel: #11 Patient ID Card
- Patient Information Booklet
- Chart Sticker
- Paper Tape Measure
- Tourniquet
- Dressing: GRIP-LOK®1 Catheter Fixation Device Valve: MicroCLAVE™2 Neutral Displacement Connector

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU Not made with natural rubber latex.

Fluid path components are non-pyrogenic.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.

All c	omponents are CE 2797 unless otherwise noted.
² A tr	ademark of ICU Medical, Inc.
'A re	gistered trademark of TIDI.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.38	847	4679	4

^{*} Priming volumes are approximate and are done without accessories.

^{††} Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.





LOT






M
















Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL059713 R01 (2023-09)















 $[\]dagger$ Flow rate values are approximate and are determined using deionized water at 100 cm head height.