



REF PR-44052-BAS (IPNXXXXXX)

5.5 Fr. 2 Lumen 40 cm catheter length .018 inch dia. spring-wire guide

# Pressure Injectable Arrowgard Blue Advance® Two-Lumen PICC

**Contents:**

- 1: Two Lumen TaperFree® Catheter with Arrowgard Blue Advance® Antimicrobial/Antithrombogenic Protection<sup>1</sup>: 5.5 Fr. (1.88 mm OD) x 40 cm, Pressure Injectable, T-Port Connector, Blue FlexTip® and Placement Wire
  - 1: GlideThru™ Peel-Away Sheath: 5.5 Fr. x 2-3/4" (7 cm) Radiopaque over 5.5 Fr. Dilator
  - 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other)
  - 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
  - 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
  - 1: Syringe: 10 mL Luer-Lock
  - 2: Dust Cap: Non-Vented
  - 1: SecondSite™ Adjustable Hub: Fastener
  - 1: SecondSite™ Adjustable Hub: Catheter Clamp
  - 1: SharpsAway® II Locking Disposal Cup
  - 1: SharpsAway® Disposal Cup
  - 1: Catheter Trimmer
  - 1: Safety Scalpel: #11
  - 1: Patient ID Card
  - 1: Patient Information Booklet
  - 2: Paper Tape Measure
  - 1: Tourniquet
  - 1: Dressing: STATLOCK®<sup>2</sup> PICC Plus Catheter Stabilization Device
- <sup>1</sup>Licensed under US Patent No. 7,329,412.  
<sup>2</sup>A registered trademark of C. R. Bard, Inc.

All components are CE 2797 unless otherwise noted.

**Rx only**  
**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU**

California Prop. 65  
**⚠ WARNING:** Cancer and Reproductive Harm.  
www.P65Warnings.ca.gov

**Not made with natural rubber latex.**  
Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).  
Fluid path components are non-pyrogenic.

**Contraindications: The Pressure Injectable Arrowgard Blue Advance antimicrobial/antithrombogenic catheter is contraindicated:**

- for patients with known hypersensitivity to chlorhexidine
- in the presence of device related infection in the intended insertion vessel or catheter pathway
- in the presence of thrombosis in the intended insertion vessel or catheter pathway



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.41	330	3850	5
Proximal (18 Ga.)	0.43	343	3880	5

\* Priming volumes are approximate and are done without accessories.  
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.  
‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.  
\*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



en cs da nl et fi fr de el it no pl pt ro ru  
sk sl es sv tr



CAS # 7440-48-4

LOT Lot Number  
YYYY-MM-DD  
YYYY-MM-DD



(01)XXXXXXXXXXXXXX  
(17)YYMMDD  
(11)YYMMDD  
(10)Lot Number

EU Authorized Representative and Importer:  
Teleflex Medical  
IDA Business and Technology Park  
Dublin Road, Athlone  
Co. Westmeath, Ireland

All components are CE 2797 unless otherwise noted.

Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

Product of xxx  
Packaged in xxx

Teleflex®  
www.teleflex.com