

Dimensions:  
Width: 8-3/4" (22,2 cm)  
Length: 11-1/8" (28,3 cm)

**PB-10001-000A**  
**PBZ-10001-000A**

T-12703-100C  
Width: 8.7598 in = 222.5mm  
Length: 10.374 in = 263.5mm

**ARROW**



**REF PR-35041-HPSM** (IPNXXXXXX)

**4** Fr. **1** Lumen **50** cm catheter length **.018** inch dia. spring-wire guide

# Pressure Injectable One-Lumen PICC with 80 cm Wire

**Contents:**

- 1: One Lumen TaperFree® Catheter: 4 Fr. (1.45 mm OD) x 50 cm, Pressure Injectable, Blue FlexTip®  
Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 4" (10 cm) Radiopaque over 4 Fr. Dilator Origin: XX XXXX XX  
XXX XXXXXXXX XXXXXXXXXX
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on Both Ends) Origin: XX  
XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Spring-Wire Guide, Marked: .018" (0.46 mm) dia. x 31-1/2" (80 cm)(Straight Soft Tip on One End -  
Straight Stiff Tip on Other) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW Origin: XX XXXX XX XXX XXXXXXXX  
XXXXXXXXXX
- 1: Syringe: 5 mL Luer-Lock Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Dust Cap: Non-Vented Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: SecondSite™ Adjustable Hub: Catheter Clamp Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: SecondSite™ Adjustable Hub: Fastener Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Patient ID Card Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Chart Sticker Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Patient Information Booklet Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Checklist/CLIP Sheet Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Paper Tape Measure Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Dressing: STATLOCK®<sup>1</sup> PICC Plus Catheter Stabilization Device Origin: XX XXXX XX XXX  
XXXXXXXX XXXXXXXXXX

<sup>1</sup>A registered trademark of C. R. Bard, Inc.

**Rx only**

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.**

[www.teleflex.com/IFU](http://www.teleflex.com/IFU)

**California Prop. 65**

**⚠ WARNING:** Cancer and Reproductive Harm.

[www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

**Not made with natural rubber latex.**

Fluid path components are non-pyrogenic.

**Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist.**



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (17 Ga.)	0.55	1013	5740	5

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

\*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.

Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

**LOT** Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXXXX  
(17)YYMMDD  
(11)YYMMDD  
(10)Lot Number

Set Product of xxx  
Packaged in xxx

**Teleflex**  
[www.teleflex.com](http://www.teleflex.com)

LBL062928 R02 (2024-12)



# Pressure Injectable One-Lumen PICC with 80 cm Wire

**ARROW**

**REF PR-35041-HPSM**

**4** Fr. **1** Lumen **50** cm catheter length **.018** inch dia. spring-wire guide