

BEF EU-05052-HPMSB

5^{Fr.}



Pressure Injectable Two-Lumen PICC

Lumen

Contents:

- 1: Two Lumen Catheter: 5 Fr. (1.81 mm OD) x 50 cm, Pressure Injectable, T-Port Connector, Blue FlexTip®, Contamination Guard and Placement Wire
- Sheath: 14 Ga. x 2-3/4" (7 cm) Radiopaque Peel-Away over 5 Fr. Dilator Spring-Wire Guide: .018" (0.46 mm) dia. x 13-1/8" (33 cm) (Straight Soft Tip on 1: Both Ends)
- Catheter: 22 Ga. x 1-3/4" (4.45 cm) Radiopaque over 25 Ga. RW Introducer 1: Needle
- 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- Injection Needle: Eclipse^{™1} 22 Ga. x 1-1/2" (3.81 cm) RW 1:
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip
- Injection Needle: Eclipse^{™1} 25 Ga. x 1" (2.54 cm) RW and 3 mL Luer-Lock 1: Syringe
- Syringe: 10 mL Luer-Lock 1:
- Syringe: 10 mL Luer-Slip 1:
- 2: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp 1:
- SharpsAway® II Locking Disposal Cup 1:
- SharpsAway® Disposal Cup 1:
- Catheter Trimmer 1:
- 1:
- Towel: 24" x 36" (61 cm x 91 cm)
- Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations 1.
- 1: Towe
- Safety Scalpel: #11 1:
- 1: Patient ID Card
- 1: Sterile Procedure Sign

Not made with natural rubber latex.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.

- 1: Patient Information Booklet
- 1: Chart Sticker
- 1: PICC Line Insertion Checklist
- 2: Paper Tape Measure
- 1: Tourniquet
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)

50 cm catheter .018 spring-wire guide

- 1: Dressing: Tegaderm^{®2} 15.5 cm x 10 cm
- Surgical Apparel: Impervious Gown 1:
- 1: Surgical Apparel: Bouffant Cap
- Surgical Apparel: Mask with Eye Shield 1:
- Dressing: STATLOCK®3 PICC Plus Catheter Stabilization Device 1:
- 1: Surgical Apparel: Mask
- Tape: Steri-Strip®2 1:
- 1: HemoHopper® Fluid Receptacle
- ¹A trademark of Becton, Dickinson and Company
- ²A registered trademark of 3M Company. 3A registered trademark of C. R. Bard, Inc

All components are CE 2797 unless otherwise noted.

de el it pl fr tr

O Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.4	356	3106	4
Proximal (18 Ga.)	0.43	396	3490	4

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



LBL063133 R02 (2023-09)