



REF PR-35052-HPHNM (IPNXXXXXX)

5 Fr. 2 Lumen 50 cm catheter length .018 inch dia. spring-wire guide

# Pressure Injectable Two-Lumen PICC with 80 cm Wire

Contents:

- 1: Two Lumen TaperFree® Catheter: 5 Fr. (1.81 mm OD) x 50 cm, Pressure Injectable, Blue FlexTip® and Contamination Guard origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: GlideThru™ Peel-Away Sheath: 5 Fr. x 4" (10 cm) Radiopaque over 5 Fr. Dilator origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Guidewire, Marked: .018" (0.46 mm) dia. x 80 cm Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil (CE 0120) origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Syringe: 10 mL Luer-Lock origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 2: Dust Cap: Non-Vented origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: SecondSite™ Adjustable Hub: Fastener origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: SecondSite™ Adjustable Hub: Catheter Clamp origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Catheter Trimmer origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Safety Scalpel: #11 origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Patient ID Card origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Chart Sticker origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Patient Information Booklet origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Checklist/CLIP Sheet origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Paper Tape Measure origin:xxxxxxxxxxxxxxxxxxxxxxxx
- 1: Dressing: STATLOCK® PICC Plus Catheter Stabilization Device origin:xxxxxxxxxxxxxxxxxxxxxxxx

¹A registered trademark of C. R. Bard, Inc.  
All components are CE 2797 unless otherwise noted.

**Rx only**  
**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU**  
 Not made with natural rubber latex.  
 Fluid path components are non-pyrogenic.  
**Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.**

| Lumen             | Priming Volume* (mL) | Gravity Flow Rate† (mL/hr) | Pump Flow Rate‡ (mL/hr) | MAX Pressure Injection Flow Rate** (mL/sec) |
|-------------------|----------------------|----------------------------|-------------------------|---|
| Distal (18 Ga.)   | 0.48                 | 488                        | 3860                    | 5   |
| Proximal (18 Ga.) | 0.5                  | 488                        | 3870                    | 5   |

\* Priming volumes are approximate and are done without accessories.  
 † Flow rate values are approximate and are determined using deionized water at 100 cm head height.  
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.  
 \*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.

EU Authorized Representative and Importer:  
 Teleflex Medical  
 IDA Business and Technology Park  
 Dublin Road, Athlone  
 Co. Westmeath, Ireland

All components are CE 2797 unless otherwise noted.

Arrow International LLC  
 Subsidiary of Teleflex Incorporated  
 3015 Carrington Mill Blvd.  
 Morrisville, NC 27560 USA

Set Product of xxx  
Packaged in xxx

Teleflex®  
www.teleflex.com

LBL063143 R03 (2024-06)



LOT LotNumber

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXX  
 (17)YYMMDD  
 (11)YYMMDD  
 (10)LotNumber

CAS # 7440-48-4

