



REF

CDA-35541-HPK1A

(IPNXXXXXX)

4 Fr.

1 Lumen

55 cm catheter length

.018 inch dia. spring-wire guide

ARROW®

ErgoPack®



Pressure Injectable One-Lumen PICC

Contents:

- 1: One Lumen TaperFree® Catheter: 4 Fr. (1.45 mm OD) x 55 cm, Pressure Injectable, T-Port Connector, Blue FlexTip®, Contamination Guard and Placement Wire Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 2-3/4" (7 cm) Radiopaque over 4 Fr. Dilator Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Injection Needle: SafetyGlide™¹ 25 Ga. x 5/8" (1.60 cm) RW Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Syringe: 10 mL Luer-Lock Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Syringe: 3 mL Luer-Lock Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Dust Cap: Non-Vented Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: SecondSite™ Adjustable Hub: Catheter Clamp Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: SecondSite™ Adjustable Hub: Fastener Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: SharpsAway® II Locking Disposal Cup Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: SharpsAway® Disposal Cup Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Catheter Trimmer Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX

- 1: Towel: 24" x 36" (61 cm x 91 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Towel Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Safety Scalpel: #11 Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Patient ID Card Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Sterile Procedure Sign Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Chart Sticker Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Patient Information Booklet Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Checklist/CLIP Sheet Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 2: Paper Tape Measure Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Tourniquet Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Surgical Apparel: Impervious Gown Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Dressing: I.V. Advanced 8.5 cm x 11.5 cm Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Surgical Apparel: Mask with Eye Shield Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Surgical Apparel: Bouffant Cap Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Surgical Apparel: Mask Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX

- XXXXXXX
- 1: Tape: Steri-Strip®² Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: HemoHopper® Fluid Receptacle Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX

¹A trademark of Becton, Dickinson and Company.

²A registered trademark of 3M Company.

All components are CE 2797 unless otherwise noted.

Not made with natural rubber latex.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.



Lumen

Distal (17 Ga.)

Priming
Volume*
(mL)

0.58

Gravity
Flow
Rate†
(mL/hr)

946

Pump
Flow Rate††
(mL/hr)

5720

MAX Pressure
Injection
Flow Rate**
(mL/sec)

5

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



CAS # 7440-48-4

LBL068110 R04 (2024-12)