



REF CDA-35563-HPK1A (IPNXXXXXX)

6 Fr. 3 Lumen 55 cm catheter length .018 inch dia. spring-wire guide



# Pressure Injectable Three-Lumen PICC

Contents:

- 1: Three Lumen Catheter: 6 Fr. (2.14 mm OD) x 55 cm, Pressure Injectable, T-Port Connector, Blue FlexTip®, Contamination Guard and Placement Wire

1: GlideThru™ Peel-Away Sheath: 6 Fr. x 2-3/4" (7 cm) Radiopaque over 6 Fr. Dilator

1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer

1: Injection Needle: SafetyGlide™ 25 Ga. x 5/8" (1.60 cm) RW

1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW

1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW

1: Syringe: 10 mL Luer-Lock

1: Syringe: 3 mL Luer-Lock

3: Dust Cap: Non-Vented

1: SecondSite™ Adjustable Hub: Fastener

1: SecondSite™ Adjustable Hub: Catheter Clamp

1: SharpsAway® II Locking Disposal Cup

1: SharpsAway® Disposal Cup

1: Catheter Trimmer

1: Towel: 24" x 36" (61 cm x 91 cm)

1: Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations

1: Towel

1: Safety Scalpel: #11

1: Patient ID Card

1: Sterile Procedure Sign

1: Chart Sticker

1: Patient Information Booklet

1: Checklist/CLIP Sheet

2: Paper Tape Measure
- 1: Tourniquet

5: Gauze Pad: 2" x 2" (5 cm x 5 cm)

5: Gauze Pad: 4" x 4" (10 cm x 10 cm)

1: Surgical Apparel: Impervious Gown

1: Dressing: I.V. Advanced 8.5 cm x 11.5 cm

1: Surgical Apparel: Mask with Eye Shield

1: Surgical Apparel: Bouffant Cap

1: Surgical Apparel: Mask


1: Tape: Steri-Strip®2

1: HemoHopper® Fluid Receptacle
- 1A trademark of Becton, Dickinson and Company.

2A registered trademark of 3M Company.
- All components are CE 2797 unless otherwise noted.

Not made with natural rubber latex.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.

 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (17 Ga.)	0.68	996	5950	6
Medial (19 Ga.)	0.49	238	2540	not rated
Proximal (19 Ga.)	0.5	257	2500	not rated

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

\*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



CAS # 7440-48-4