
PB-10000-850 8.5 X 10.5



Pressure Injectable One-Lumen PICC with 80 cm Wire

Contents:

One Lumen TaperFree® Catheter: 4 Fr. (1.45 mm OD) x 55 cm, Pressure 1. Injectable, Blue FlexTip® and Contamination Guard GlideThru™ Peel-Away Sheath: 4 Fr. x 4" (10 cm) Radiopaque over 4 Fr. Dilator Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) 1: (Straight Soft Tip on One End - Straight Stiff Tip on Other) Guidewire, Marked: .018" (0.46 mm) dia. x 80 cm Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil (CE 0120) Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW Syringe: 10 mL Luer-Lock Dust Cap: Non-Vented 1. SecondSite™ Adjustable Hub: Catheter Clamp SecondSite™ Adjustable Hub: Fastener 1. Catheter Trimmer Safety Scalpel: #11 Patient ID Card 1. Chart Sticker Patient Information Booklet Checklist/CLIP Sheet Paper Tape Measure 1: Dressing: STATLOCK®1 PICC Plus Catheter Stabilization Device 1: ¹A registered trademark of C. R. Bard, Inc. All components are CE 2797 unless otherwise noted.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (17 Ga.)	0.58	946	5720	5

Not made with natural rubber latex

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

++ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.

//CAS # 7440-48-4

LBL069184 R01 (2025-01)

