

REF PR-34063-HPHNM





Pressure Injectable Three-Lumen PICC with 80 cm Wire

Contents:

- 1: Three Lumen TaperFree® Catheter: 6 Fr. (2.14 mm OD) x 40 cm, Pressure Injectable, Blue FlexTip® and Contamination Guard
- 1: GlideThru™ Peel-Away Sheath: 6 Fr. x 4" (10 cm) Radiopaque over 6 Fr. Dilator
- Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on Both Ends)
- 1: Guidewire, Marked: .018" (0.46 mm) dia. x 80 cm Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil (CE 0120)
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- 3: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite[™] Adjustable Hub: Catheter Clamp
- 1: Catheter Trimmer
- 1: Safety Scalpel: #11
- 1: Patient ID Card
- 1: Chart Sticker
- 1: Patient Information Booklet
- 1: Checklist/CLIP Sheet
- 1: Paper Tape Measure
- 1: Dressing: STATLOCK®1 PICC Plus Catheter Stabilization Device
- ¹A registered trademark of C. R. Bard, Inc.

Not made with natural rubber latex.

All components are CE 2797 unless otherwise noted.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (17 Ga.)	0.55	1187	6800	6
Medial (19 Ga.)	0.44	343	2820	not rated
Proximal (19 Ga.)	0.42	363	2920	not rated

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



contraindications exist.



Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no

LBL069368 R01 (2023-09)