Seal/Weld Area

Dimensions:
Width: 272 mm
Length: 365 mm

Dimensions:
Width: 10.7087"
Length: 14.3701"

T-20000-100, 103, 104, 150

ARROW

REF NA-11242-CDC (IPNXXXXXX)

9 Fr. distal lumen | **2** Lumen | **11.5** cm device length | **.035** inch dia. spring-wire guide

MAC™ Two-Lumen Central Venous Access for use with 7.5 - 8 Fr. Catheters

Contents:

1: Two-Lumen Central Venous Access Device: 115 mm Radiopaque Polyurethane with 9 Fr. Distal Lumen, Integral Hemostasis Valve, Extension Line Clamps Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Injection Needle: SafetyGlide™[1] 25 Ga. x 1" (2.54 cm) Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Pressure Transduction Probe Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Injection Needle: SafetyGlide™[1] 23 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Syringe: 3 mL Luer-Lock Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Tissue Dilator Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®[2] One-Step Solution with Hi-Lite Orange™ Tint Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Obturator: 8 Fr. Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Cath-Gard® Catheter Contamination Shield: 80 cm with TwistLock™ Adapter Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Dust Cap: Vented Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
2: Dust Cap: Non-Vented Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: SharpsAway® II Locking Disposal Cup Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: SharpsAway® Disposal Cup Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Drape: 41" x 55" with 4" fenestration Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Needle Holder Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Safety Scalpel: #11 Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Stopcock: 4-way High-Flow Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Medication Label: 1% Lidocaine Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
2: Gauze Pad: 2" x 2" (5 cm x 5 cm) Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
5: Gauze Pad: 4" x 4" (10 cm x 10 cm) Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Suture: 3-0 Silk with Curved Needle Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX
1: Tubing±: 20" (50 cm) Extension Origin: XX XXXX XX XXX XXXXXXX XXXXXXXXXX

[1]A trademark of Becton, Dickinson and Company.
[2]A registered trademark of CareFusion or one of its subsidiaries.

Rx only
Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.
www.teleflex.com/IFU
California Prop. 65
⚠ **WARNING:** Cancer and Reproductive Harm.
www.P65Warnings.ca.gov
±Precaution: Contains Phthalate: DEHP.
Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.
Not made with natural rubber latex.
Store between 20 – 25°C (68 – 77°F). Avoid freezing and excessive heat above 40°C (104°F).
Sterilized using ethylene oxide unless otherwise indicated in the contents list.
Fluid path components are non-pyrogenic.
Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (9 Fr.) no catheter	1.8	31440
Distal (9 Fr.) with 7 Fr. catheter	1.45	16280
Distal (9 Fr.) with 8 Fr. catheter	1.15	9040
Proximal (12 Ga.) no catheter	0.6	8970

* Priming volumes are approximate and are done without accessories.
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

MD | STERILE EO | LATEX | 20°C (68°F) 25°C (77°F) | PHT ± DEHP

LOT LotNumber
YYYY-MM-DD
YYYY-MM-DD
UDI
(01)XXXXXXXXXXXXXX
(17)YYMMDD
(11)YYMMDD
(10)LotNumber

www.teleflex.com/IFU

Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Set Product of xxx
Packaged in xxx

Teleflex®
www.teleflex.com

LBL069457 R03 (2025-07)

REF NA-11242-CDC
9 Fr. distal lumen | 2 Lumen | 11.5 cm device length | .035 inch dia. spring-wire guide
MAC™ Two-Lumen Central Venous Access for use with 7.5 - 8 Fr. Catheters