



REF PR-45541-HPHNL (IPNXXXXXX)

4.5 Fr. **1** Lumen **55** cm catheter length **.018** inch dia. spring-wire guide

Pressure Injectable Arrowg+ard Blue Advance® One-Lumen PICC with 130 cm Hydrophilic-Coated Nitinol Guidewire

Contents:

- 1: One Lumen TaperFree® Catheter with Arrowgard Blue Advance® Antimicrobial/Antithrombogenic Protection: 4.5 Fr. (1.58 mm OD) x 55 cm Pressure Injectable and Contamination Guard
- 1: GlideThru™ Peel-Away Sheath: 4.5 Fr. x 4" (10 cm) Radiopaque over 4.5 Fr. Dilator
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other)
- 1: Guidewire, Marked: .018" (0.46 mm) dia. x 130 cm Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- 1: Dust Cap: Non-Vented
- 1: Catheter Trimmer
- 1: Safety Scalpel: #11
- 1: Patient ID Card
- 1: Patient Information Booklet
- 1: Paper Tape Measure
- 1: Dressing: STATLOCK® 2 PICC Plus Catheter Stabilization Device

¹Licensed under US Patent No. 7,329,412 and 11,219,706.

²A registered trademark of C. R. Bard, Inc.

All components are CE 2797 unless otherwise noted.

Rx only

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm.

WARNING: Cancer and
www.P65Warnings.ca.gov

###MR Safe


Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).

Fluid path components are non-pyrogenic.

Contraindications: The Pressure Injectable Arrow®+ard Blue Advance antimicrobial/antithrombogenic catheter is contraindicated:

- for patients with known hypersensitivity to chlorhexidine
- in the presence of device related infection in the intended insertion vessel or catheter pathway
- in the presence of thrombosis in the intended insertion vessel or catheter pathway

 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Pink (17 Ga.)	0.66	1090	5380	5

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.




LOT LotNumber

 YYYY-MM-DD




(01)XXXXXXXXXXXXXXXX
(17)YYMMDD
(11)YYMMDD
(10)LotNumber



 EU Authorized
Representative and
Importer:

 Teleflex Medical
IDA Business and
Technology Park
Dublin Road, Athlone
Co. Westmeath, Ireland

All components are CE 2797 unless otherwise noted.

 Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Product of xxx
Packaged in xxx

Teleflex®
www.teleflex.com

LBL070047 R01 (2024-05)