
Product of XXX Packaged in XXX









(IPNXXXXXXX)

## Arrowg+ard Blue Plus® Three-Lumen CVC

- Arrowg+ard Blue Plus®1 Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 16 cm
- Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One 1. End - "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW 1.
- Pressure Transduction Probe
- Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL
- Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm 1.
- 3: Dust Cap: Non-Vented
- SecondSite™ Adjustable Hub: Fastener 1:
- SecondSite™ Adjustable Hub: Catheter Clamp

<sup>1</sup>Licensed under US Patent No. 7,329,412.

All components are CE 2797 unless otherwise noted.

Rx only Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU Not made with natural rubber latex. Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F). Fluid path components are non-pyrogenic.

Contraindications: The Arrowg+ard Blue Plus antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa



Distal (16 Ga.) 0.38 2490	Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Proximal (18 Ga.) 0.39 1328	Medial (18 Gá.)	0.37	1162	

ow rate values are approximate and are determined using deionized water at 100 cm head height.





EU Authorized Representative and Importer:


Teleflex Medical IDA Business and Technology Park Dublin Road, Athlone 2797 Co. Westmeath, Ireland

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL074534 R00 (2024-03)

OT LotNumber



YY-MM-DD



(01)XXXXXXXXXXXXXXX (17)YYMMDD (11)YYMMDD

(10)LotNumber





catheter

\$\,\text{032} \text{ spring-wire auide}



