



REF EU-00740-45

(IPNXXXXXX)



Insertion Set for use with Arrow® 4 Fr. PICC

Contents:

- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 2-3/4" (7 cm) Radiopaque over 4 Fr. Dilator
Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft
Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer Origin: XX XXXX XX
XXX XXXXXXXX XXXXXXXXXX
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW Origin: XX XXXX XX XXX
XXXXXXXX XXXXXXXXXX
- 1: SharpsAway® II Locking Disposal Cup Origin: XX XXXX XX XXX XXXXXXXX
XXXXXXXXXXXX
- 1: Safety Scalpel: #11 Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX

All components are CE 2797 unless otherwise noted.

Rx only Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm.
www.P65Warnings.ca.gov

Not made with natural rubber latex.

Fluid path components are non-pyrogenic.



EU Authorized
Representative and
Importer:



Teleflex Medical
IDA Business and
Technology Park
Dublin Road, Athlone
Co. Westmeath, Ireland

All components are CE 2797 unless
otherwise noted.

Set Product of xxx
Packaged in xxx

LOT Lot Number



YYYY-MM-DD



YYYY-MM-DD



(01)XXXXXXXXXXXXXX
(17)YYMMDD
(11)YYMMDD
(10)Lot Number



Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Teleflex®
www.teleflex.com

LBL075402 R00 (2025-07)

