



7 Fr. **3** Lumen **16** cm catheter length **.032** inch dia. spring-wire guide

Contents:


- 1: Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 16 cm
 - 1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
 - 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
 - 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW
 - 1: Pressure Transduction Probe
 - 1: Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL
 - 1: Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
 - 3: Dust Cap: Non-Vented
 - 1: SecondSite™ Adjustable Hub: Fastener
 - 1: SecondSite™ Adjustable Hub: Catheter Clamp
- All components are CE 2797 unless otherwise noted.**

Rx only
Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU
California Prop. 65
 ⚠ **WARNING:** Cancer and Reproductive Harm.
www.P65Warnings.ca.gov
Not made with natural rubber latex.
 Fluid path components are non-pyrogenic.

 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (16 Ga.)	0.4	2947
Medial (18 Ga.)	0.4	1552
Proximal (18 Ga.)	0.4	1702

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

en be bg hr cs da nl et fi fr de el hu is
it lv lt no pl pt ro ru sr sk sl es sv tr uk



EU Authorized
Representative and
Importer:



P Teleflex Medical
IDA Business and
Technology Park
Dublin Road, Athlone
Co. Westmeath, Ireland

All components are CE 2797 unless otherwise noted.



 Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

LOT Lot Number

YYYY-MM-DD



YYYY-MM-DD



(01)XXXXXXXXXXXXXXXX
(17)YYMMDD
(11)YYMMDD
(10)Lot Number

Product of xxx
Packaged in xxx

Teleflex®
www.teleflex.com

LBL076568 R00 (2025-07)



REF CS-12703-E

7 Fr. **3** Lumen

16 cm catheter length

.032 inch dia.
spring-wire
guide



Three-Lumen CVC