



IDLX-35541-HPK

(IPNXXXXXX)

4 Fr.

1 Lumen

55 cm
catheter
length

.018 inch dia.
spring-wire
guide

ARROW®

Pressure Injectable One-Lumen PICC



Contents:

- | | |
|--|---|
| 1: One Lumen TaperFree® Catheter, †††: 4 Fr. (1.45 mm OD) x 55 cm, Pressure Injectable, Stiffening Stylet (0.36 mm dia. x 75 cm with straight distal tip) and T-Port Connector | 2: Paper Tape Measure |
| 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 2-3/4" (7 cm) Radiopaque over 4 Fr. Dilator | 1: Tourniquet |
| 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer | 5: Gauze Pad: 2" x 2" (5 cm x 5 cm) |
| 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW | 10: Gauze Pad: 4" x 4" (10 cm x 10 cm) |
| 1: Injection Needle: Eclipse™1 25 Ga. x 1" (2.54 cm) | 1: Surgical Apparel: Impervious Gown |
| 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW | 1: Dressing: I.V. Advanced 8.5 cm x 11.5 cm |
| 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW | 1: Transducer Cover with Gel and Rubber Bands |
| 1: Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip | 1: Surgical Apparel: Mask with Eye Shield |
| 1: Syringe: 10 mL Luer-Lock | 1: Dressing: STATLOCK®2 PICC Plus Catheter Stabilization Device |
| 1: Syringe: 3 mL Luer-Lock | 1: Surgical Apparel: Bouffant Cap |
| 1: Dust Cap: Non-Vented | 1: Surgical Apparel: Mask |
| 1: SecondSite™ Adjustable Hub: Catheter Clamp | 1: Tape: Steri-Strip®3 |
| 1: SecondSite™ Adjustable Hub: Fastener | 1: Valve: MicroCLAVE™4 Neutral Displacement Connector |
| 1: SharpsAway® Disposal Cup | |
| 1: SharpsAway® II Locking Disposal Cup | |
| 1: Catheter Trimmer | |
| 1: Towel: 24" x 36" (61 cm x 91 cm) | |
| 1: Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations | |
| 1: Towel | |
| 1: Safety Scalpel: #11 | |
| 1: Patient ID Card | |
| 1: Sterile Procedure Sign | |
| 1: Patient Information Booklet | |

1A trademark of Becton, Dickinson and Company.

2A registered trademark of C. R. Bard, Inc.

3A registered trademark of 3M Company.

4A trademark of ICU Medical, Inc.

All components are CE 2797 unless otherwise noted.

***MR Safe

Not made with natural rubber latex.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.



Lumen

Priming
Volume*
(mL)

Gravity
Flow
Rate†
(mL/hr)

Pump
Flow Rate††
(mL/hr)

MAX Pressure
Injection
Flow Rate**
(mL/sec)

Pink (17 Ga.)

0.63

970

5300

5

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



CAS # 7440-48-4

†††

