



REF PR-05042 (IPNXXXXXX)

4 Fr. 2 Lumen 50 cm catheter length .018 inch dia. spring-wire guide

Two-Lumen PICC

Contents:

- 1: Two Lumen Catheter: 4 Fr. (1.45 mm OD) x 50 cm, Blue FlexTip®
- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 4" (10 cm) Radiopaque over 4 Fr. Dilator
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other)
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 5 mL Luer-Slip
- 2: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: Patient ID Card
- 1: Chart Sticker
- 1: Patient Information Booklet
- 1: Paper Tape Measure
- 1: Dressing: STATLOCK®¹ PICC Plus Catheter Stabilization Device

¹A registered trademark of C. R. Bard, Inc.

All components are CE 2797 unless otherwise noted.

Rx only

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm. www.P65Warnings.ca.gov

Not made with natural rubber latex.

Fluid path components are non-pyrogenic.


Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.



CAS # 7440-48-4

en

es

 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)
Distal (22 Ga.)	0.27	165	1446
Proximal (22 Ga.)	0.33	139	1529

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.



EU Authorized Representative and Importer:



Teleflex Medical
IDA Business and
Technology Park
Dublin Road, Athlone
Co. Westmeath, Ireland



All components are CE 2797 unless otherwise noted.



Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Product of xxx
Packaged in xxx

LOT Lot Number



YYYY-MM-DD



YYYY-MM-DD



(01)XXXXXXXXXXXXXXX
(17)YYMMDD
(11)YYMMDD
(10)Lot Number

Teleflex®
www.teleflex.com

LBL077246 R00 (2025-09)



4 Fr. 2 Lumen 50 cm catheter length .018 inch dia. spring-wire guide

REF PR-05042

Two-Lumen PICC

