







55 cm catheter length

## Pressure Injectable Three-Lumen PICC pre-loaded with Arrow® NaviCurve™ Stylet



## Contents:

- Three Lumen TaperFree® Catheter‡‡: 6 Fr. (2.14 mm OD) x 55 cm, Pressure
- Injectable, Arrow® NaviCurve™ Stylet and T-Port Connector
- GlideThru™ Peel-Away Sheath: 6 Fr. x 2-3/4" (7 cm) Radiopaque over 6 Fr.
- Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm)
- (Straight Soft Tip on One End Straight Stiff Tip on Other) with Arrow Advancer
- Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- Injection Needle: Eclipse™1 25 Ga. x 1" (2.54 cm)
- Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip
- Syringe: 10 mL Luer-Lock
- Syringe: 3 mL Luer-Lock
- Dust Cap: Non-Vented
- SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp
- SharpsAway® Disposal Cup
- SharpsAway® II Locking Disposal Cup
- Catheter Trimmer
- Towel: 24" x 36" (61 cm x 91 cm)

Not made with natural rubber latex.

- Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations
- Towel

###MR Safe

- Safety Scalpel: #11
- Sterile Procedure Sign Patient Information Booklet
- 2: Paper Tape Measure

contraindications exist.

- 1: Tourniquet
- 1: Foam Electrodes (3 per pack)
- Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- Surgical Apparel: Impervious Gown
- Dressing: GRIP-LOK®2 Catheter Fixation Device
- Dressing: I.V. Advanced 8.5 cm x 11.5 cm
- Transducer Cover with Gel and Rubber Bands
- Remote Cover
- Surgical Apparel: Mask with Eye Shield
- Surgical Apparel: Bouffant Cap
- Surgical Apparel: Mask
- Tape: Steri-Strip®3
- Valve: MicroCLAVE™4 Neutral Displacement Connector
- 1: T-Piece Cover
- <sup>1</sup>A trademark of Becton, Dickinson and Company
- <sup>2</sup>A registered trademark of TIDI
- <sup>3</sup>A registered trademark of 3M Company
- <sup>4</sup>A trademark of ICU Medical, Inc.
- All components are CE 2797 unless otherwise noted.




















<sup>\*</sup> Priming volumes are approximate and are done without accessories.





Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no

LBL078824 R00 (2025-03)

<sup>†</sup> Flow rate values are approximate and are determined using deionized water at 100 cm head height.

<sup>††</sup> Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

<sup>\*\*</sup> Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.