

(IPNXXXXXX)

30 cm catheter length

.032 inch dia. spring-wire guide

## Arrowg+ard Blue® Three-Lumen CVC

- Arrowg+ard Blue® Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 30 cm Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One End "J" Tip on Other)
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW
  Pressure Transduction Probe

- Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
- Dust Cap: Non-Vented
- SecondSite™ Adjustable Hub: Fastener SecondSite™ Adjustable Hub: Catheter Clamp
- All components are CE 2797 unless otherwise noted.

Rx	onl	y
----	-----	---

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

Fluid path components are non-pyrogenic.

Contraindications: The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (16 Ga.)	0.6	2025	
Medial (18 Ga.)	0.5	921	
Proximal (18 Ga.)	0.5	1004	





<u>/</u>...∕CAS # 7440-48-4

EU Authorized Representative and Importer:

Teleflex Medical IDA Business and Technology Park
Dublin Road, Athlone
7797 Co. Westmeath, Ireland

All components are CE 2797 unless otherwise noted.

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA

Product of xxx Packaged in xxx

**Neleflex**®

LBL079154 R00 (2025-08)





LOT Lot Number

√ YYYY-MM-DD







/// | YYYY-MM-DD
















(01)XXXXXXXXXXXXXXX (17)YYMMDD (11)YYMMDD









<sup>\*</sup> Priming volumes are approximate and are done without accessories. † Flow rate values are approximate and are determined using deionized water at 100 cm head height.