





(IPNXXXXXXX



Pressure Injectable One-Lumen Midline



Contents:

- 1: TaperFree® Catheter±±±: 4 Fr.(1.45 mm OD) x 20 cm Pressure Injectable Radiopaque Polyurethane with Extension Line Clamp and Contamination Guard
- GlideThru™ Peel-Away Sheath: 4 Fr. x 2-3/4" (7 cm) Radiopaque over 4
- Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer
- Injection Needle: Eclipse™1 25 Ga. x 1" (2.54 cm)
- Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- Syringe: 5 mL Luer-Lock
- 10 mL Luer-Lock Pre-Filled Saline Syringe
- 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®2 One-Step Solution 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®2 One-Step Solution
- with Hi-Lite Orange™ Tint
- Dust Cap: Non-Vented
- SharpsAway® II Locking Disposal Cup
- SharpsAway® Disposal Cup
- Catheter Trimmer
- Towel: 24" x 36" (61 cm x 91 cm)
- Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations
- Safety Scalpel: #11
- 1: Sterile Procedure Sign

- 1: Overbed Sign
- 1: Medication Label: 1% Lidocaine
- Patient Information Booklet
- Paper Tape Measure
- 10: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Surgical Apparel: Impervious Gown
- Dressing: Tegaderm®3 CHG 3-1/2" x 4-1/2" (8.5 cm x 11.5 cm)
- Transducer Cover with Gel and Rubber Bands
- Surgical Apparel: Mask with Eye Shield
- Dressing: STATLOCK®4 PICC Plus Catheter Stabilization Device
- Surgical Apparel: Bouffant Cap
- Surgical Apparel: Mask
- 1: Tape: Steri-Strip®3
- ¹A trademark of Becton, Dickinson and Company,
- ²A registered trademark of CareFusion or one of its subsidiaries.
- ³A registered trademark of 3M Company.
- ⁴A registered trademark of C. R. Bard, Inc.

Not made with natural rubber latex.

Store between 20 - 25°C (68 - 77°F).

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Yellow (18 Ga.)	0.37	1880	9324	5

^{*} Priming volumes are approximate and are done without accessories.





LBL080075 R00 (2025-08)

[†] Flow rate values are approximate and are determined using deionized water at 100 cm head

^{††} Pump flow rates are determined at maximum pump pressure of 10 psig and represent

^{**} Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with 152 cm administration set tubing.