



REF PS-01652-NM
(IPNXXXXXX)

ARROW®

Two-Lumen PICC

Contents:

- Two Lumen Catheter: 16 Ga. (1.76 mm OD) x 50 cm, Blue FlexTip®, Contamination Guard
- Peel-Away Catheter: 14 Ga. x 2-1/2" (6.35 cm) Radiopaque over 15 Ga. TW Introducer Needle and 3 mL Luer-Slip Syringe
- Dust Cap: Non-Vented
- Clamp: Extension Line
- SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp
- Drape: 18" x 26" (46 cm x 66 cm) with 3" (7.6 cm) fenestration
- Dressing: STATLOCK®¹ PICC Plus Catheter Stabilization Device

¹A registered trademark of C. R. Bard, Inc.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

Not made with natural rubber latex. Fluid path components are non-pyrogenic.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)
Distal (18 Ga.)	0.42	449	3444
Proximal (20 Ga.)	0.43	178	2015

* Priming volumes are approximate and are done without accessories.
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.
†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.



Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Product of xxx
Packaged in xxx

LOT Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXX
(17)YYMMDD
(11)YYMMDD
(10)Lot Number

LBL082216 R00 (2026-06)



REF PS-01652-NM
(IPNXXXXXX)

ARROW®

Two-Lumen PICC

Contents:

- Two Lumen Catheter: 16 Ga. (1.76 mm OD) x 50 cm, Blue FlexTip®, Contamination Guard
- Peel-Away Catheter: 14 Ga. x 2-1/2" (6.35 cm) Radiopaque over 15 Ga. TW Introducer Needle and 3 mL Luer-Slip Syringe
- Dust Cap: Non-Vented
- Clamp: Extension Line
- SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp
- Drape: 18" x 26" (46 cm x 66 cm) with 3" (7.6 cm) fenestration
- Dressing: STATLOCK®¹ PICC Plus Catheter Stabilization Device

¹A registered trademark of C. R. Bard, Inc.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

Not made with natural rubber latex. Fluid path components are non-pyrogenic.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)
Distal (18 Ga.)	0.42	449	3444
Proximal (20 Ga.)	0.43	178	2015

* Priming volumes are approximate and are done without accessories.
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.
†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.



Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Product of xxx
Packaged in xxx

LOT Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXX
(17)YYMMDD
(11)YYMMDD
(10)Lot Number

LBL082216 R00 (2026-06)



16 Ga. 2 Lumen

50 cm catheter length

REF PS-01652-NM
Two-Lumen PICC



16 Ga. 2 Lumen

50 cm catheter length

REF PS-01652-NM
Two-Lumen PICC