



ARROW® Midline Kit

REF CDC-02051-MK1A

5 Fr. | 20 cm catheter length | .018 inch dia. spring-wire guide

ErgoPack®

Contents:

- 1: Midline Catheter, Trimmable: 5 Fr. x 20 cm Radiopaque Polyurethane with T-Port Connector, Integral Extension Line, Extension Line Clamps, Integral Suture Wing, Contamination Guard and Placement Wire
- 1: GlideThru™ Peel-Away Sheath: 5 Fr. x 2-3/4" (7 cm) Radiopaque over 5 Fr. Dilator
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 13-1/8" (33 cm) (Straight Soft Tip on Both Ends)
- 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- 1: Injection Needle: SafetyGlide™ 25 Ga. x 5/8" (1.60 cm) RW
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- 1: Syringe: 3 mL Luer-Lock
- 1: Skin Protectant Prep Pad (non-sterile solution)
- 2: 10 mL Luer-Lock Pre-Filled Saline Syringe
- 1: 5 mL 1% Lidocaine HCl Solution
- 1: 3 mL Applicator 2% CHG and 70% IPA Chloraprep®² One-Step Solution with Hi-Lite Orange™ Tint (non-sterile solution)

- 1: Dust Cap: Non-Vented
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Catheter Trimmer
- 1: Towel: 24" x 36" (61 cm x 91 cm)
- 1: Drape: 60" x 76" (152 cm x 193 cm)
- 1: Drape: 36" x 41" (91 cm x 104 cm) with 3" x 5" (8 cm x 13 cm) fenestration
- 1: Towel
- 1: Safety Scalpel: #11
- 1: Sterile Procedure Sign
- 1: Medication Label: 1% Lidocaine
- 1: Checklist/CLIP Sheet
- 2: Paper Tape Measure
- 1: Tourniquet
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm®³ 15.5 cm x 10 cm

- 1: Surgical Apparel: Impervious Gown
- 1: Dressing: BIOPATCH®⁴
- 1: Dressing: STATLOCK®⁵ Catheter Stabilization Device
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Surgical Apparel: Mask
- 1: Tape: Steri-Strip®³
- 1: HemoHopper® Fluid Receptacle

¹A trademark of Becton, Dickinson and Company.
²A registered trademark of CareFusion or one of its subsidiaries.
³A registered trademark of 3M Company.
⁴A registered trademark of Johnson & Johnson Corporation.
⁵A registered trademark of C. R. Bard, Inc.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.


Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.



Not made with natural rubber latex



Store between 20 - 25°C (68 - 77°F).



Lumen	Priming Volume* (mL)	Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)
Distal (18 Ga.)	0.35	3060	11290

* Priming volumes are approximate and are done without accessories.
† Flow rates were determined using room temperature water, 100 cm head height and represent approximate flow capabilities.
‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.