



Arrow®

Pressure Injectable PICC Kit



REF **EU-05541-HPMSB**

4 Fr. | **55** cm catheter length | **.018** inch dia. spring-wire guide

Contents:

- 1: Indwelling Catheter: 4 Fr. x 55 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, T-Port Connector, Extension Line Clamps, Contamination Guard and Placement Wire
- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 2-3/4" (7 cm) Radiopaque over 4 Fr. Dilator
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 13-1/8" (33 cm) (Straight Soft Tip on Both Ends)
- 1: Catheter: 22 Ga. x 1-3/4" (4.45 cm) Radiopaque over 25 Ga. RW Introducer Needle
- 1: Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip
- 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- 1: Injection Needle: Eclipse™ 22 Ga. x 1-1/2" (3.81 cm) RW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Injection Needle: Eclipse™ 25 Ga. x 1" (2.54 cm) RW and 3 mL Luer-Lock Syringe
- 1: Syringe: 10 mL Luer-Lock
- 1: Syringe: 10 mL Luer-Slip
- 1: Skin Protectant Prep Pad (non-sterile solution)
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Catheter Trimmer
- 1: Towel: 24" x 36" (61 cm x 91 cm)
- 1: Drape: 60" x 76" (152 cm x 193 cm)
- 1: Drape: 36" x 41" (91 cm x 104 cm) with 3" x 5" (8 cm x 13 cm) fenestration
- 1: Towel

- 1: Safety Scalpel: #11
- 1: Sterile Procedure Sign
- 1: Flow Rate Information Card with Injection Log
- 1: Patient ID Card
- 1: Chart Sticker
- 1: Patient Information Booklet
- 1: PICC Line Insertion Checklist
- 2: Paper Tape Measure
- 1: Tourniquet
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm™ 15.5 cm x 10 cm
- 1: Surgical Apparel: Impervious Gown
- 1: Dressing: STATLOCK™ Catheter Stabilization Device
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Surgical Apparel: Mask
- 1: Tape: Steri-Strip™
- 1: HemoHopper® Fluid Receptacle

¹A trademark of Becton, Dickinson and Company.

²A registered trademark of 3M Company.

³A registered trademark of C. R. Bard, Inc.

All components are CE 0086 unless otherwise noted.


Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist.



Not made with natural rubber latex

- nl** Kit met Arrow PICC voor hogedrukinjectie
- fr** Kit de CCIP pour injection sous pression Arrow
- de** Arrow PEZK-Kit für Druckinjektionen
- el** Kit PICC για έγχυση υπό πίεση Arrow
- it** Kit con PICC iniettabile a pressione Arrow
- pl** Nadający się do wstrzykiwania pod ciśnieniem zestaw PICC Arrow
- pt** Kit de cateter central de inserção periférica Arrow para injeção pressurizada
- ru** Комплект катетера Arrow PICC для введения веществ под давлением
- es** Kit de catéter central de inserción periférica para inyección a presión Arrow
- sv** Arrow tryckinjicerbart PICC-set
- tr** Arrow Basınçla Enjekte Edilebilir PICC Kit



Lumen	Priming Volume* (mL)	Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.50	660	4075	4

* Priming volumes are approximate and are done without accessories.

† Flow rates were determined using room temperature water, 100 cm head height and represent approximate flow capabilities.

‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.