



REF MP-04301-CJ

Single Lumen Central Venous Catheterization Set

16^{Ga.} (1.7^{mm}) | 1^{Lumen} | 20^{cm} catheter length | .018^{inch dia.} spring-wire guide


Contents:

- | | |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------|
| 1: Indwelling Catheter: 16 Ga. x 8" (20 cm) Radiopaque Polyurethane with Blue FlexTip®, Integral Extension Line, Extension Line Clamp, Integral Suture Wing and Stylet | 1: Syringe: 2 mL Luer-Slip |
| 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 23-5/8" (60 cm) (Straight Stiff Tip on One End - "J" Tip on Other) with Arrow Advancer and hub | 1: Tissue Dilator: 6 Fr. x 6.5 cm |
| 1: Catheter: 20 Ga. (0.84 mm) x 1-3/4" (4.45 cm) Radiopaque over 22 Ga. (0.72 mm) TW Introducer Needle | 1: Dust Cap: Non-Vented |
| 1: Introducer Needle: Echogenic 22 Ga. (0.72 mm) x 2-3/4" (7 cm) XTW | 1: SecondSite™ Adjustable Hub: Fastener |
| 1: Introducer Needle: Echogenic 22 Ga. (0.72 mm) x 1-3/8" (3.5 cm) XTW | 1: SecondSite™ Adjustable Hub: Catheter Clamp |
| | 2: Guidewire Introducer |
| | 1: Y-hub |
| | 3: Towel: 13" x 18" (33 cm x 46 cm) |

Rx only. Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

Not made with natural rubber latex.

Fluid path components are non-pyrogenic.



Catheter Diameter: 16 Ga., 1.7 mm

Priming Volume: 0.43 mL

Priming volumes are approximate and are done without accessories.

Flow Rate: 2900 mL/hr

Based on 100 cm head height.



STERILE EO



LOT



Arrow International, Inc.
Subsidiary of Teleflex Incorporated
2400 Bernville Road
Reading, PA 19605 USA

Teleflex®
www.teleflex.com

SZ-04301-125B (09/19)



REF MP-04301-CJ

16^{Ga.} (1.7^{mm}) | 1^{Lumen} | 20^{cm} catheter length | .018^{inch dia.} spring-wire guide

Single Lumen Central Venous Catheterization Set